

TAMILNADU MEDICAL SERVICES CORPORATION LIMITED

**Tender no. PKG4/ICB/TNUHP/JICA/C2/END/TNMSC/ENGG/2020
Dt. 04.02.2020** for supply and installation of **Upper GI Video Endoscope
Entire Set, Video Duodenoscope with ERCP Accessories entire set,
Video Colonoscope Entire Set, Endoscopic Washer Disinfector, Video
Bronchoscope (Flexible), Ultraslim Video Gastroscope under JICA**

Tender inviting officer, authority, designation and address:

The General Manager (Equipment),
Tamil Nadu Medical Services Corporation Ltd.,
Egmore, Chennai – 600 008.

Date of Invitation of bids: 17.02.2020

1. Name of the Equipment: Upper GI Video Endoscope Entire Set (*Lot/Contract no.1*)

2. Make: Fujifilm, Japan

Model:

- a. *Video processor:* VP-3500HD
- b. *Light Source:* XL-4450 (300W Xenon)
- c. *Video Endoscope:* EG-600 WR V2
- d. *Water jet flushing pump:* JW-2
- e. *Medical grade monitor:* LG-27HJ710S

3. Qty: 13 nos.

4. Delivery places:

Medical College Hospitals: (2 nos. each)

1. Nagercoil
2. Salem
3. Thanjavur
4. Trichy
5. Vellore
6. Tirunelveli
7. Thoothukudi – 1 no. only

5. No. of bids received: 3 (three) only

6. Details of bidders participated with their quoted model:

S.No.	Name and address of the firm	Equipment Details
1	M/s Fujifilm India Pvt. Ltd., 6 th Floor, Prestige Palladium Bayan, Door nos: 129 to 140, Greems Road, Chennai – 600 006.	Make: Fujifilm, Japan Model: a. <i>Video processor:</i> VP-3500HD b. <i>Light Source:</i> XL-4450 (300W Xenon) c. <i>Video Endoscope:</i> EG-600 WR V2 d. <i>Water jet flushing pump:</i> JW-2 e. <i>Medical grade monitor:</i> LG-27HJ710S
2	M/s Olympus Medical Systems India Pvt. Ltd., 101-First Floor, Capital Tower, 180-Kodambakkam High Road, Nungambakkam, Chennai – 600 034.	Make: Olympus, Japan Model: a. <i>Video processor:</i> Optera CV-170 b. <i>Light Source:</i> LED Light Source (integrated) c. <i>Video Endoscope:</i> Optera GIF-H170 d. <i>Water jet flushing pump:</i> OFP-2 e. <i>Medical grade monitor:</i> Sony LMD-2735
3	M/s Mitra Medical Services LLP, B-226, OKHLA Industrial Area, Phase –I, New Delhi – 110 020.	Make: Mitra, India (Trademark: Ottomed) Model: a. <i>Video processor:</i> OVP-SEII b. <i>Light Source:</i> Built-in, LED at tip c. <i>Video Endoscope:</i> OVG-mB2H d. <i>Water jet flushing pump:</i> mEP-II e. <i>Medical grade monitor:</i> Sony LMD- 2735 MD

7. Shortlisted Bidders:

- a) M/s Fujifilm India Pvt. Ltd.,
- b) M/s Olympus Medical Systems India Pvt. Ltd.,

8. Bidder whose bid was rejected and reasons:

M/s Mitra Medical Services LLP

Technical Points:

(iii) Video Duodenoscope:

- a) Scope shall be equipped of inbuilt identification memory chip, scope's model & serial no. display at screen.
- b) Observation Range: 4-60 mm or better
- c) Shall have G-lock or dual lock mechanism for better treatment and reduce procedure time.

(v) Medical Grade Monitor:

- a) Resolution 1920 x 1200
- b) Aspect ratio 16:10
- c) Viewing angle 178 degree

"Shall have European CE Marked or US FDA"- **ISO and ICMED certification only available.**

Commercial Points:

Not pertaining to minimum requirement of tender

- a) Form EXP-3: Production Capacity
- b) Form FIN-2: Average Annual Turnover
- c) Form EXP-2: Specific Experience

9. Price details:

Name of the firm		M/s Olympus Medical Systems India Pvt. Ltd.,	M/s Fujifilm India Pvt. Ltd.,
Qty		13	13
Unit price w/o tax (INR)		13,34,078.00	15,67,999.00
Total price w/o tax (INR)		1,73,43,014.00	2,03,83,987.00
CAMC (INR)	4th year	****	47,039.00
	5th year	****	47,039.00
	6th year	****	47,039.00
	7th year	****	47,039.00
	8th year	****	47,039.00
	9th year	****	47,039.00
	10th year	****	47,039.00
Total CAMC Discounted Value (excl. GST)(INR)		----	27,29,278.04
Evaluated price with CAMC at NPV rate of 8% per annum for 13 nos. (INR)		----	2,31,13,265.04
Position		----	L1

**** **Note:** M/s Olympus Medical Systems had enclosed CAMC rate of **Lot/Contract no. 3 (Video Colonoscope)** instead of **Upper GI Endoscope (Lot/Contract no.1)**, which was intimated to them in email dt. 25.01.2021. Again in email dt. 04.02.2021, they were informed that have failed to submit/enclose the price bid for CAMC pertaining to Lot/Contract no. 1, Upper GI Video Endoscope Entire set, instead they have enclosed another copy of price bid for CAMC pertaining to Lot/Contract no.3, Video Colonoscope Entire set.

Hence their price bid for Lot/ Contract no.1 is treated as incomplete and thereby considered as Non-Responsive.

10. Award details:

The Purchase order Ref.

PKG4/ICB/TNUHP/JICA/C2/END/TNMSC/ENGG/2020 dt. 02.03.2021

was placed on responsive L1 bidder, M/s Fujifilm India Pvt. Ltd., at a total value of **INR 2,03,83,987/- (excl. GST) for 13 Nos.**

1. Name of the Equipment: Video Duodenoscope with ERCP Accessories Entire set

(Lot/Contract no.2)

2. Make: Fujifilm, Japan

Model:

- a. *Video processor:* VP-3500HD
- b. *Light Source:* XL-4450 (300W Xenon)
- c. *Video Duodenoscope:* ED-580XT
- d. *Medical grade monitor:* LG-27HJ710S

3. Qty: 10 nos.

4. Delivery places:

Medical College Hospitals: (1 no. each)

- a) Kanyakumari
- b) Trichy
- c) Vellore
- d) Thoothukudi
- e) Salem (2 nos.)

f) Thanjavur (2 nos.) and

g) Tirunelveli (2 nos.)

5. No. of bids received: 3 (three) only

6. Details of bidders participated with their quoted model:

S.No.	Name and address of the firm	Equipment Details
1	M/s Fujifilm India Pvt. Ltd., 6 th Floor, Prestige Palladium Bayan, Door nos: 129 to 140, Greems Road, Chennai – 600 006.	Make: Fujifilm, Japan Model: <i>a. Video processor:</i> VP-3500HD <i>b. Light Source:</i> XL-4450 (300W Xenon) <i>c. Video Duodenoscope:</i> ED-580XT <i>d. Medical grade monitor:</i> LG-27HJ710S
2	M/s Olympus Medical Systems India Pvt. Ltd., 101-First Floor, Capital Tower, 180-Kodambakkam High Road, Nungambakkam, Chennai – 600 034.	Make: Olympus, Japan Model: <i>a. Video processor:</i> EVIS EXERA III CV-190 <i>b. Light Source:</i> EVIS EXERA III CLV-190 (300W Xenon) <i>c. Video Duodenoscope:</i> EVIS EXERA III TJF-Q190V <i>d. Medical grade monitor:</i> Sony LMD-2735
3	M/s Mitra Medical Services LLP, B-226, OKHLA Industrial Area, Phase -I, New Delhi – 110 020.	Make: Mitra, India (Trademark: Ottomed) Model: <i>a)Video processor:</i> OVP-SEII <i>b)Light Source:</i> Built-in, LED at tip <i>c)Video Duodenoscope:</i> OVD-2H <i>d)Medical grade monitor:</i> Sony LMD-2735 MD

7. Shortlisted Bidders:

a) M/s Fujifilm India Pvt. Ltd.,

b) M/s Olympus Medical Systems India Pvt. Ltd.,

8. Bidder whose bid was rejected and reasons:

M/s Mitra Medical Services LLP

Technical Points:

(iii) Video Duodenoscope:

- a) Scope shall be equipped of inbuilt identification memory chip, scope's model & serial no. display at screen.
- b) Observation Range: 4-60 mm or better
- c) Shall have G-lock or dual lock mechanism for better treatment and reduce procedure time.

(v) Medical Grade Monitor:

- a) Resolution 1920 x 1200
- b) Aspect ratio 16:10
- c) Viewing angle 178 degree

"Shall have European CE Marked or US FDA"- **ISO and ICMED certification only available.**

Commercial Points:

Not pertaining to minimum requirement of tender

- a) Form EXP-3: Production Capacity
- b) Form FIN-2: Average Annual Turnover
- c) Form EXP-2: Specific Experience

9. Price details:

Name of the firm	M/s Fujifilm India Pvt. Ltd.,	M/s Olympus Medical Systems India Pvt. Ltd.,	
Qty	10	10	
Unit price w/o tax (INR)	15,70,166.00	39,34,966.00	
Total price w/o tax (INR)	1,57,01,660.00	3,93,49,660.00	
CAMC (INR)	4th year	47,104.00	1,27,056.00
	5th year	47,104.00	1,27,056.00
	6th year	47,104.00	1,27,056.00
	7th year	47,104.00	1,27,056.00
	8th year	47,104.00	1,27,056.00
	9th year	47,104.00	1,27,056.00
	10th year	47,104.00	1,27,056.00
Evaluated price with CAMC at NPV rate of 8% per annum for 10 nos. (INR)	1,78,04,005.73	4,50,20,423.39	
Position	L1	L2	

10. Award details:

The Purchase order **Ref.**

PKG4/ICB/TNUHP/JICA/C2/END/TNMSC/ENGG/2020 dt. 02.03.2021

was placed on responsive L1 bidder, M/s Fujifilm India Pvt. Ltd., at a total value of **INR 1,57,01,660/- (excl. GST) for 10 Nos.**

1. Name of the Equipment: Video Colonoscope Entire Set (*Lot/Contract no. 3*)

2. Make: Fujifilm, Japan

Model:

- a. *Video processor:* VP-3500HD
- b. *Light Source:* XL-4450 (300W Xenon)
- c. *Video Colonoscope:* EC-530WL3
- d. *CO2 Insufflator:* GW-100
- e. *Medical grade monitor:* LG-27HJ710S

3. Qty: 12 nos.

4. Delivery places:

Medical College Hospitals: (2 nos. each)

- a) Kanyakumari
- b) Salem
- c) Thanjavur
- d) Trichy
- e) Tirunelveli
- f) Vellore (1 no.) and
- g) Thoothukudi (1 no.)

5. No. of bids received: 3 (three) only

6. Details of bidders participated with their quoted model:

S.No.	Name and address of the firm	Equipment Details
1	M/s Fujifilm India Pvt. Ltd., 6 th Floor, Prestige Palladium Bayan, Door nos: 129 to 140, Greems Road, Chennai – 600 006.	Make: Fujifilm, Japan Model: <i>a. Video processor:</i> VP-3500HD <i>b. Light Source:</i> XL-4450 (300W Xenon) <i>c. Video Colonoscope:</i> EC-530WL3 <i>d. CO2 Insufflator:</i> GW-100 <i>e. Medical grade monitor:</i> LG-27HJ710S

S.No.	Name and address of the firm	Equipment Details
2	M/s Olympus Medical Systems India Pvt. Ltd., 101-First Floor, Capital Tower, 180-Kodambakkam High Road, Nungambakkam, Chennai – 600 034.	Make: Olympus, Japan Model: a. <u>Video processor:</u> Optera CV-170 b. <u>Light Source:</u> LED Light Source (integrated) c. <u>Video Colonoscope:</u> Optera CF-H170L d. <u>CO2 Insufflator:</u> UCR e. <u>Medical grade monitor:</u> Sony LMD-2735
3	M/s Mitra Medical Services LLP, B-226, OKHLA Industrial Area, Phase -I, New Delhi – 110 020.	Make: Mitra, India (Trademark: Ottomed) Model: a) <u>Video processor:</u> OVP-SEII b) <u>Light Source:</u> Built-in, LED at tip c) <u>Video Colonoscope:</u> OVC-mB2H d) <u>CO2 Insufflator:</u> m-EC-III e) <u>Medical grade monitor:</u> Sony LMD- 2735 MD

7. Shortlisted Bidders:

- a) M/s Fujifilm India Pvt. Ltd.,
- b) M/s Olympus Medical Systems India Pvt. Ltd.,

8. Bidder whose bid was rejected and reasons:

M/s Mitra Medical Services LLP

Technical Points:

(iii) Video Colonoscope:

- a) Scope shall be equipped of inbuilt identification memory chip, scope's model & serial no. display at screen.
- b) Observation Range: 2-100 mm including normal and closest.
- c) Minimum visible distance of 2 mm closer from distal end.

(iv) Technical Specifications of CO2 Insufflator:

a) Dimension 125(W) x 300(D) x 150(H)

(v) Medical Grade Monitor:

a) Resolution 1920 x 1200

b) Aspect ratio 16:10

c) Viewing angle 178 degree

(vii) "Shall have European CE Marked or US FDA"- **ISO and ICMED certification only available.**

Commercial Points:

Not pertaining to minimum requirement of tender

a) Form EXP-3: Production Capacity:

b) Form FIN-2: Average Annual Turnover:

c) Form EXP-2: Specific Experience

9. Price details:

Name of the firm	M/s Fujifilm India Pvt. Ltd.,	M/s Olympus Medical Systems India Pvt. Ltd.,
Qty	12	12
Unit price w/o tax (INR)	14,02,666.00	14,36,970.00
Total price w/o tax (INR)	1,68,31,992.00	1,72,43,640.00
CAMC (INR)	4th year	88,852.00
	5th year	88,852.00
	6th year	88,852.00
	7th year	88,852.00
	8th year	88,852.00
	9th year	88,852.00
10th year	88,852.00	
Evaluated price with CAMC at NPV rate of 8% per annum for 12 nos. (INR)	1,90,54,665.60	2,20,02,410.96
Position	L1	L2

10. Award details:

The Purchase order Ref.

PKG4/ICB/TNUHP/JICA/C2/END/TNMSC/ENGG/2020 dt. 02.03.2021

was placed on responsive L1 bidder, M/s Fujifilm India Pvt. Ltd., at a total value of **INR 1,68,31,992/- (excl. GST) for 12 Nos.**

1. Name of the Equipment: Endoscopic Washer Disinfector (*Lot/Contract no.5*)

2. Make: Huons Medicare Co. Ltd., Korea

Model: Well-2

3. Qty: 13 nos.

4. Delivery places:

Medical College Hospitals: (2 nos. each)

- a) Kanyakumari
- b) Salem
- c) Thanjavur
- d) Trichy
- e) Vellore
- f) Tirunelveli and
- g) Thoothukudi (1 no.)

5. No. of bids received: 3 (three) only

6. Details of bidders participated with their quoted model:

S.No.	Name and address of the firm	Equipment Details
1	M/s Fujifilm India Pvt. Ltd., 6 th Floor, Prestige Palladium Bayan, Door nos: 129 to 140, Greems Road, Chennai – 600 006.	Make: Huons Medicare Co. Ltd., Korea Model: Well-2
2	M/s. South India Surgical Co. Ltd., #117/65, Wallajah Road, Chennai – 600 002.	Make: Medipia Co. Ltd., Korea Model: Endo Clean 2000
3	M/s Mitra Medical Services LLP, B-226, OKHLA Industrial Area, Phase –I, New Delhi – 110 020.	Make: Mitra, India (Trademark: Ottomed) Model: OEW-2

7. Shortlisted Bidder: M/s Fujifilm India Pvt. Ltd.,

8. List of bidders whose bids were rejected and reasons:

(A) M/s Mitra Medical Services LLP

Technical Points:

(iii) Disinfection Method:

a) It shall have tank capacity-18 lit.

(iv) Air Shower Method:

a) Air force to channels-timing shall be adjustable to 10 to 30 sec.

(v) Alcohol Wash Method:

a. Alcohol purging is available for flushing the working channels.

b. Tank Capacity shall be of 1 lit.

(v) Machine Dimensions(mm):

a. It shall have 630 (W) x 7000 (D) x 980 mm (H)

b. Weight: 90 kilogram

- Optional foot pedal operation shall be available, not needed to open with hands.

(vii) "Shall have European CE Marked or US FDA"- ISO and ICMED certification only available.

Commercial Points:

Not pertaining to minimum requirement of tender

- Form EXP-3: Production Capacity
- Form EXP-2: Specific Experience

(B) M/s South India Surgical Co. Ltd., Chennai

Commercial Points:

Not pertaining to minimum requirement of tender

For them as a Non-Manufacturer:

- Form EXP-1: Not furnished
- Form EXP-2: Specific Experience- Not furnished
- Form EXP-3: Production and Sales Capacity- Not furnished.
- Confirmation that all facilities exist with the bidder (or manufacturer, as applicable) in its factory for inspection and testing and these can

be accessed by the Purchaser or its representative for inspection-
Not furnished.

For their Manufacturer:

- a) Form FIN-1: Not Furnished
- b) Form FIN-2: Not Furnished
- c) Form FIR-1: Not Furnished
- d) Form FIR-2: Not Furnished
- e) Form EXP-1: Not Furnished
- f) Form EXP-2: Not Furnished
- g) Form EXP-3: Not Furnished

9. Price details:

Name of the firm	M/s Fujifilm India Pvt. Ltd.,	
Qty	13	
Unit price w/o tax (INR)	6,75,000.00	
Total price w/o tax (INR)	87,75,000.00	
CAMC (INR)	4th year	14,580.00
	5th year	14,580.00
	6th year	14,580.00
	7th year	14,580.00
	8th year	14,580.00
	9th year	14,580.00
	10th year	14,580.00
Evaluated price with CAMC at NPV rate of 8% per annum for 13 nos. (INR)	96,20,954.93	
Position	L1	

10. Award details:

The Purchase order **Ref.**

PKG4/ICB/TNUHP/JICA/C2/END/TNMSC/ENGG/2020 dt. 02.03.2021

was placed on responsive L1 bidder, M/s Fujifilm India Pvt. Ltd., at a total value of **INR 87,75,000/- (excl. GST) for 13 Nos.**

1. Name of the Equipment: *Video Bronchoscope (Flexible) (Lot/Contract no.6)*

2. Make: Fujifilm, Japan

Model:

- a) Video processor: VP-3500HD
- b) Light Source: XL-4450 (300W Xenon)
- c) Video Bronchoscope: EB-530T

d)Medical grade monitor: LG-27HJ710S

3. Qty: 2 nos.

4. Delivery places:

Medical College Hospitals: (2 nos. each)

a) Tirunelveli

b) Salem

5. No. of bids received: 3 (three) only

6. Details of bidders participated with their quoted model:

S.No.	Name and address of the firm	Equipment Details
1	M/s Fujifilm India Pvt. Ltd., 6 th Floor, Prestige Palladium Bayan, Door nos: 129 to 140, Greams Road, Chennai – 600 006.	Make: Fujifilm, Japan Model: <i>a)Video processor:</i> VP-3500HD <i>b)Light Source:</i> XL-4450 (300W Xenon) <i>c) Video Bronchoscope:</i> EB-530T <i>d)Medical grade monitor:</i> LG-27HJ710S
2	M/s Olympus Medical Systems India Pvt. Ltd., 101-First Floor, Capital Tower, 180-Kodambakkam High Road, Nungambakkam, Chennai – 600 034.	Make: Olympus, Japan Model: <i>a)Video processor:</i> CV-170 <i>b) Light Source:</i> LED Light Source (integrated) <i>c)Video Bronchoscope:</i> BF-1T150 ACTERA <i>d) Medical grade monitor:</i> Sony LMD-2735
3	M/s Mitra Medical Services LLP, B-226, OKHLA Industrial Area, Phase –I, New Delhi – 110 020.	Make: Mitra, India (Trademark: Ottomed) Model: <i>a)Video processor:</i> OVP-SEII

S.No.	Name and address of the firm	Equipment Details
		<i>b)Light Source:</i> Built-in, LED at tip <i>c) Video Colonoscope:</i> OCB-IN, OLT <i>d)Medical grade monitor:</i> Sony LMD- 2735 MD

7. Shortlisted Bidders:

- a) M/s Fujifilm India Pvt. Ltd.,
- b) M/s Olympus Medical Systems India Pvt. Ltd.,

8. Bidder whose bid was rejected and reasons:

M/s Mitra Medical Services LLP

Technical points:

(i) Video Processor:

- a) Digital to Digital Recording option shall be available for both still and moving images.
- b) Must have USB flash drive compatibility for storing images.
- c) Picture in Picture/ POP display option
- d) Anti-blur function and NR function to get sharpest and clearest images for review and documentation.
- e) Shall be upgradeable for Ultrasound Gastroscope and Balloon Enteroscope in future.

(ii) Light Source:

- a) Shall be controllable in three steps to feed the air/water as required.
- b) Forced air cooling shall be available.

(v) Medical Grade Monitor:

- a) Aspect ratio 16:10
- b) Display Colors 16.8 million

(vii) "Shall have European CE Marked or US FDA"- **ISO and ICMED certification only available.**

Commercial Points:

Not pertaining to minimum requirement of tender

Form EXP-2: Specific Experience

9. Price details:

Name of the firm	M/s Fujifilm India Pvt. Ltd.,	M/s Olympus Medical Systems India Pvt. Ltd.,
Qty	2	2
Unit price w/o tax (INR)	12,96,666.00	12,05,495.00
Total price w/o tax (INR)	25,93,332.00	24,10,990.00
CAMC (INR)	4th year	59,492.00
	5th year	59,492.00
	6th year	59,492.00
	7th year	59,492.00
	8th year	59,492.00
	9th year	59,492.00
	10th year	59,492.00
Evaluated price with CAMC at NPV rate of 8% per annum for 2 nos. (INR)	29,32,535.20	29,42,039.39
Position	L1	L2

10. Award details:

The Purchase order Ref.

PKG4/ICB/TNUHP/JICA/C2/END/TNMSC/ENGG/2020 dt. 02.03.2021

was placed on responsive L1 bidder, M/s Fujifilm India Pvt. Ltd., at a total value of **INR 25,93,332/- (excl. GST) for 2 Nos.**

1. Name of the Equipment: *Ultraslim Video Gastroscope (Lot/Contract no.7)*

2. Make: Fujifilm, Japan

Model:

- a) Video processor: VP-3500HD
- b) Light Source: XL-4450 (300W Xenon)
- c) Video Gastroscope: EG-580NW2
- d) Medical grade monitor: LG-27HJ710S

3. Qty: 6 nos.

4. Delivery places:

Medical College Hospitals: (1 no. each)

- a) Kanyakumari,
- b) Salem,
- c) Thanjavur,
- d) Trichy,
- e) Tirunelveli and
- f) Thoothukudi.

5. No. of bids received: 3 (three) only

6. Details of bidders participated with their quoted model:

S.No.	Name and address of the firm	Equipment Details
1	M/s Fujifilm India Pvt. Ltd., 6 th Floor, Prestige Palladium Bayan, Door nos: 129 to 140, Greems Road, Chennai – 600 006.	Make: Fujifilm, Japan Model: a) Video processor: VP-3500HD b) Light Source: XL-4450 (300W Xenon) c) Video Gastroscope: EG-580NW2 d) Medical grade monitor: LG-27HJ710S
2	M/s Olympus Medical Systems India Pvt. Ltd., 101-First Floor, Capital Tower, 180-Kodambakkam High Road, Nungambakkam, Chennai – 600 034.	Make: Olympus, Japan Model: a) Video processor: CV-170 b) Light Source: LED Light Source (integrated) c) Video Gastroscope: GIF-XP170N d) Medical grade monitor: Sony LMD-2735
3	M/s Mitra Medical Services LLP, B-226, OKHLA Industrial Area, Phase –I, New Delhi – 110 020.	Make: Mitra, India (Trademark: Ottomed) Model: a) Video processor: OVP-SEII b) Light Source: Built-in, LED at tip c) Video Gastroscope: OVG-2HUS d) Medical grade monitor: Not Offered

7. Shortlisted Bidders:

- a) M/s Fujifilm India Pvt. Ltd.,
- b) M/s Olympus Medical Systems India Pvt. Ltd.,

8. Bidder whose bid was rejected and reasons:

M/s Mitra Medical Services LLP

Technical points:

(iii) Video Gastroscope:

a) Scope shall be equipped of inbuilt identification memory chip, scope's model & serial no. display at screen.

- Distal End Outer Diameter: 5-6.5 mm
- Insertion tube Outer Diameter: 5.5-6.5 mm or better

1. Video Processor-Scope is an add-on compatible with OVP-SEII Processor.
2. Light Source- Scope is an add-on compatible with OVP-SEII Processor.
3. Medical Grade Monitor- Not offered.
4. Endoscopic Cart- Not offered.

(vii) "Shall have European CE Marked or US FDA"- **ISO and ICMED certification only available.**

Commercial Points:

Not pertaining to minimum requirement of tender

- a) Form EXP-3: Production and Sales Capacity: Not furnished
- b) Form EXP-2: Specific Experience

9. Price details:

Name of the firm		M/s Fujifilm India Pvt. Ltd.,	M/s Olympus Medical Systems India Pvt. Ltd.,
Qty		6	6
Unit price w/o tax (INR)		14,57,666.00	13,87,578.00
Total price w/o tax (INR)		87,45,996.00	83,25,468.00
CAMC (INR)	4th year	43,730.00	71,623.00
	5th year	43,730.00	71,623.00
	6th year	43,730.00	71,623.00
	7th year	43,730.00	71,623.00
	8th year	43,730.00	71,623.00
	9th year	43,730.00	71,623.00
	10th year	43,730.00	71,623.00
Evaluated price with CAMC at NPV rate of 8% per annum for 6 nos. (INR)		99,17,050.42	1,02,43,474.64
Position		L1	L2

10. Award details:

The Purchase order **Ref.**

PKG4/ICB/TNUHP/JICA/C2/END/TNMSC/ENGG/2020 dt. 02.03.2021
was placed on responsive L1 bidder, M/s Fujifilm India Pvt. Ltd., at a total value
of **INR 87,45,996/- (excl. GST) for 6 Nos.**