

27.11.2021

**TAMILNADU MEDICAL SERVICES CORPORATION LTD.,**

**TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND INSTALLATION OF  
FULLY AUTOMATED HAEME ANALYZER FOR VETERINARY USE AND BIO-  
CHEMICAL AUTO ANALYZER FOR VETERINARY USE AT VARIOUS  
VETERINARY INSTITUTIONS**

**Tender No.603/ANL/AHD/TNMSC/ENGG/2020, Dt.09.10.2020**

**Corrigendum**

1. The following clarification are issued:-

<b>Sl. No.</b>	<b>Tender document reference</b>	<b>Points raised</b>	<b>Clarification</b>
2.	Page No.54 to 56  Section VI: Technical Specification  3. Specification for Bio-chemical Auto Analyzer Veterinary Use  <b>And refer corrigendum issued on 08.09.2021.</b>  <b>Accessories</b>  4. ton or more air conditioner for optimal operational environment maintenance for functioning of equipment.	Request to delete this requirement.	No change. Hence, published specification prevails.
5.	Page No.54 to 56  Section VI: Technical Specification  6. Specification for Bio-	Requested to clarify this requirement.	Bidder should provide a stand to keep the accessories like computer and printer.

Sl. No.	Tender document reference	Points raised	Clarification
	chemical Auto Analyzer Veterinary Use  <b>And refer corrigendum issued on 08.09.2021.</b>  <b>Accessories</b>  Storage platform for equipment and its accessoires		

The tender due date is extended as follows:-

Sale of Bid document up to : 09.12.2021

Date of submission of bid : 10.12.2021 up to 11.00 AM

Date of opening of bids : 10.12.2021 up to 12.00 Noon

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

**Sd/-**  
**General Manager (E)**