

Date: 16.04.2020

SUPPLY AND INSTALLATION OF FACIAL NERVE MONITOR, HIGH SPEED MICRO DRILL, NASAL ENDOSCOPY UNIT, FESS HIGH END HD CAMERA AND MICRODEBRIDER FOR MEDICAL COLLEGE HOSPITALS AND DISTRICT HEAD QUARTERS HOSPITAL AT VARIOUS DISTRICTS OF TAMIL NADU

PKG23/LCB/TNUHP/TNMSC/ENGG/2019, DT:25.10.2019

Result of Evaluation

S. No	Name of the Bidders	Equipment Details	Responsiveness		Reason for Rejection
			Technical Evaluation Report	Commercial Evaluation Report	
	1. FACIAL NERVE MONITOR – 6 Nos.				
1	M/s Dynamed Equipments, Chennai	Make: Inomed, Germany Model: C2 Nerve Monitor	Non-Responsive	Responsive	As per the demonstration report not meeting these technical specification points: 1. Shall have Real-time continuous monitoring with 'Automatic Periodic Stimulation' TM electrode to allow surgeon to take immediate corrective action to prevent potential injury. 2. Shall have Incrementing probe to adjust stimulus level and print or save by the surgeon from the sterile area. 3. Must have inbuilt nerve and electrode placement guide. 4. Shall have touch screen control for ease of use. 5. Must have an option to combine an electric drill with stimulation which warns surgeon of bur's proximity to facial nerve. 6. Must have Microscope overlay. 7. Must have function to trend amplitude and latency in real time to help prevent false negative and/or accidental nerve damage.
	M/s The Advanced Medical Systems Inc., Chennai	Make: Medtronic, Netherlands Model: NIM-Neuro 3.0	Responsive	Responsive	-----
	2.HIGH SPEED MICRO DRILL – 12 Nos.				
2	M/s Pridex Medicare Pvt. Ltd., Pune	Make: Bien Air Surgery, Switzerland Model: Osseoduo	Responsive	Responsive	-----
	M/s. Indian Orthocare	Make: Stryker, India			Not meeting these technical specification points:

S. No	Name of the Bidders	Equipment Details	Responsiveness		Reason for Rejection
			Technical Evaluation Report	Commercial Evaluation Report	
	Surgicals, Chennai	Model: RemB (Core 2 Console with Integral IRR)	Non-Responsive	Responsive	1. Hand piece should adapt any common type of bur.
	M/s Medilife Technologies, New Delhi	Make: Nouvag AG, Switzerland Model: High Surg 30	Non-Responsive	Responsive	Not meeting these technical specification points: 1. Hand piece should adapt any common type of bur.
	M/s. Smith & Nephew Healthcare Pvt. Ltd, Mumbai	Make: Spiggle & Theis, GmbH, Germany Model: Speedsurg Drill System	Non-Responsive	Responsive	Not meeting these technical specification points: 1. Hand piece should adapt any common type of bur.
	M/s The Advanced Medical Systems Inc., Chennai	Make: Medtronic, Netherlands Model: Indigo	Non-Responsive	Responsive	Not meeting these technical specification points: 1. Hand piece should adapt any common type of bur.

3. NASAL ENDOSCOPY UNIT – 10 Nos.

3	M/s Olympus Medical Systems India Pvt. Ltd., Chennai	Make: Olympus, Japan Model: 1. Endoscope: 4 mm Ultra Sinuscopes 2. 150 watts LED light source: CLL-V1 LED Light Source	Responsive	Responsive	-----
	M/s. Vishal Surgical Equipment Co. Pvt. Ltd., Chennai	Make: Karlstorz, Germany Model: 1. Endoscope: Hopkins II 2. 150 watts LED light source: LED Nova 150, Cold Light Fountain	Responsive	Responsive	-----
	M/s. Indian Orthocare	Make: Stryker, India			Not meeting these technical specification points:

S. No	Name of the Bidders	Equipment Details	Responsiveness		Reason for Rejection
			Technical Evaluation Report	Commercial Evaluation Report	
.	Surgicals, Chennai	Model: <u>1.Endoscope:</u> Ideal Eyes (Cat no. 502104010/30/45/70) <u>2. 150 watts LED light source:</u> <u>Option 1:</u> RK180W LED Light source; Kare IOS 4mmx10 ft fiberoptic cable <u>Option 2:</u> Stryker Precision LED Light Source; Stryker 5mmx10 ft fiberoptic cable	Non-Responsive	Responsive	As per the demonstration report, the pictures viewed through the telescopes 0°, 30°,45° and 70° are not clear. Hence not recommended.
4. FESS HIGH END HD CAMERA – 4 Nos.					
4	M/s. Indian Orthocare Surgicals, Chennai	Make: Stryker, India Model: 1. Three Chip High definition Camera System: 1488 HD Camera 2. High Definition Medical Grade Monitor: Stryker Vision Pro 26" LED Display 3. Xenon/LED Light Source with Fiber optic cable: <u>Option 1:</u> RK180W LED Light source; Kare IOS 4mmx10 ft fiberoptic cable <u>Option2:</u> Stryker Precision LED Light Source; Stryker 5mmx10 ft fiberoptic cable 4. Endoscopic Trolley: IOS Endoscopy Trolley	Non-Responsive	Responsive	As per the demonstration report not meeting these technical specification points: 1. <u>USB Port for Capturing FULL HD Videos/ HD Stills in External USB drive and direct interface of USB Printer to facilitate direct printouts:</u> This feature is not available. 2. <u>Modular design: Digital FULL HD camera module should be compatible for use with video flexible GI endoscopes:</u> This feature is not available. Only HD Camera available, which is an obsolete technology.

S. No	Name of the Bidders	Equipment Details	Responsiveness		Reason for Rejection
			Technical Evaluation Report	Commercial Evaluation Report	
	M/s Olympus Medical Systems India Pvt. Ltd., Chennai	Make: Olympus, Japan Model: <u>1. Three Chip High definition Camera System:</u> OTV-S200 CH-S200-XZ-EB (4K) <u>2. High Definition Medical Grade Monitor:</u> LMD-2735 <u>3. Xenon/LED Light Source with Fiber optic cable:</u> Integrated Light Source (LED) <u>4. Endoscopic Trolley:</u> A10000000410	Responsive	Responsive	-----
	M/s. Vishal Surgical Equipment Co. Pvt. Ltd., Chennai	Make: Karlstorz, Germany Model: <u>1. Three Chip High definition Camera System:</u> Image 1S Connect <u>2. High Definition Medical Grade Monitor:</u> TM 220 27" Full HD Monitor <u>3. Xenon/LED Light Source with Fiber optic cable:</u> Cold Light Fountain Xenon Nova 300 <u>4. Endoscopic Trolley:</u> A10000000410	Responsive	Responsive	-----
5.MICRODEBRIDER – 5 Nos.					
	M/s. Indian Orthocare	Make: Stryker, India	Non-Responsive	Responsive	As per the demonstration report not meeting these technical specification points:

S. No	Name of the Bidders	Equipment Details	Responsiveness		Reason for Rejection
			Technical Evaluation Report	Commercial Evaluation Report	
.	Surgicals, Chennai	Model: Core 2 Console with Integral IRR			1. <u>The Micro Debrider Hand piece should be lightest in weight and should be weighing less than 370g including the cord for the comfort of the surgeon for doing the surgeries in longer duration:</u> Hand piece in the quoted model is bulky to use during surgeries.
	M/s Medilife Technologies, New Delhi	Make: Nouvag AG, Switzerland Model: High Surg 30	Non-Responsive	Responsive	As per the demonstration report not meeting these technical specification points: 1. Non-Debrider Hand piece with DC brushless motor alone shown for demo. Debrider component not furnished during demo.
	M/s. Smith & Nephew Healthcare Pvt. Ltd, Mumbai	Make: Spiggle & Theis, GmbH, Germany Model: Speedsurg Shaver System	Non-Responsive	Responsive	As per the demonstration report not meeting these technical specification points: 1. <u>The Micro Debrider Hand piece should be lightest in weight and should be weighing less than 370g including the cord for the comfort of the surgeon for doing the surgeries in longer duration:</u> Hand piece in the quoted model is bulky to use during surgeries.
	M/s The Advanced Medical Systems Inc., Chennai	Make: Medtronic, Netherlands Model: Straight Shot M5	Responsive	Responsive	-----

Sd/-
General Manager (E)