

Dt.28.08.2020

EVALUATION REPORT

Tender no.557/OTT/DME-RGGGH/TNMSC/ENGG/2020, Dt.09.01.2020

Tender for supply and installation of Fully Motorized C-Arm Compatible OT Table to Institute of Vascular Surgery, Rajiv Gandhi Govt. General Hospital, Chennai

Sch.no.1 Fully Motorized C-Arm Compatible OT Table – 1 no.

Sl. no	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. BET Medical (P) Ltd, Make: Hotborn Medical Equipment Co. Ltd, China, Model: DC 2000J	Responsive	Responsive	-----
2	M/s. Staas Bio-med Engineering Pvt. Ltd, Make: Staas, Model: BOON	Non-Responsive	Responsive	1. The quoted model BOON of make Staas, doesn't satisfy the table movement (as it provided the longitudinal travel only 80cm & lateral travel 20cm, against the tender requirement of 82cm & 24cm. 2. Also, the catalogue doesn't specify the trendelenburg movement & degree of lateral tilt. Hence the quoted model is not-recommended.