

**Dt. 07.11.2020**

**EVALUATION REPORT**

**Tender for Supply and installation of Video Bronchoscope to  
Department of Tuberculosis and Chest Disease, Government Medical  
College and ESIC Hospital, Coimbatore**

**Tender No. 536/SCOPE/DME-CBE/TNMSC/ENGG/2019, Dt.  
11.10.2019.**

-----

<b>Sl. No</b>	<b>Name of the Bidder</b>	<b>Technical Evaluation report</b>	<b>Commercial evaluation report</b>	<b>Reason for rejection</b>
<b>1</b>	<b>M/s. Olympus Medical Systems India Pvt. Ltd.,</b> Model : BF type 1T150:CV-170: LMD 2110 Sony 21" Medical Grade Monitor Make : Olympus	<b>Responsive</b>	<b>Responsive</b>	<b>-----</b>
<b>2</b>	<b>M/s. Fujifilm India Private Limited</b> Model : EP 6000 with EB 580T Make : Fujifilm Corporation, Japan	<b>Non-Responsive</b>	<b>Responsive</b>	The quoted model does not meet the following technical specification :  Should be compatible with semi rigid thoracoscope.