

**TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND INSTALLATION
OF SPIROMETER WITH EXTERNAL PRINTER TO VARIOUS GOVT.
INSTITUTIONS**

TENDER NO.483/SPIRO/NHM/TNMSC/ENGG/2019, DT.14.06.2019

Result of Evaluation

Sl. No	Bidders Quoted	Model / Make	Responsive bidders		Reason for rejection
			Technical evaluation report	Commercial evaluation report	
1.	M/s. Peso Medicare Pvt. Ltd.	Make: Medical Equipment , Europe, Germany Model: Smart Pft USB	Responsive	Responsive	-----
2.	M/s. Solokrafts Industries	Make: MeHow, China Model: LA310	Non-Responsive	Responsive	1) As per specialist report the quoted model is not meeting the parameter MIP & MEP. Hence their quoted model is not recommended.
3.	M/s. Madras Surgicals and Chemicals	Make: ndd Medizinitechni AG, Switzerland Model: Easy on PC	Non-Responsive	Responsive	1) As per specialist report the quoted model is not meeting the Flow measurement: Pneumotach technology, hence their model is not recommended.
4.	M/s. TBS INDIA Telematic and Biomedical Services Pvt. Ltd.	Make: M/s. Cosmed SRL Model: Pony FX Mp MEP	Non-Responsive	Non-Responsive	1) As per specialist report the quoted model is not meeting the Flow measurement: Pneumotach technology, hence their model is not recommended. 2) Despite given chance the bidder did not furnish clarification / documents for the following points:

Sl. No	Bidders Quoted	Model / Make	Responsive bidders		Reason for rejection
			Technical evaluation report	Commercial evaluation report	
					<p>a. Documentary evidence for satisfactory performance of the item quoted, to meet Qualification Criteria</p> <p>b. Affidavit conforming the performance statement.</p> <p>Hence not recommended</p>
5.	M/s. Elektronik Lab India Pvt. Ltd.	Make:SIBELMED, Spain Model:Datospir aira Disposable	Non-Responsive	Responsive	<p>1) As per specialist report the quoted model is not meeting the Volume range: 0 to ± 14 L per second hence their model is not recommended.</p> <p>2) Despite given chance the bidder did not furnish clarification / documents for the following points:</p> <p>a. Documentary evidence for satisfactory performance of the item quoted, to meet Qualification Criteria</p> <p>b. Affidavit conforming the performance statement.</p> <p>c. Manufacturing capacity.</p> <p>d. Manufacturer's Authorization Form – Section XI</p> <p>Hence not recommended</p>