EVALUATION REPORT

Tender for fixing rate contract for supply and Installation of Wax bath, Continuous Passive Motion (Computerized) and Height Adjusting Tilting Table to trauma care centres at various Government Institutions

Tender ref. No. 464/PHY/TRC/TNMSC/ENGG/2019, Dt. 17.03.2020.

Sch	Name of the Equipment and Bidder	Responsiveness of bidders		
Scn No		Demo report	Commercial evaluation report	Reason for rejection
1	Wax bath M/s. Bharat Medical Systems Make : Own	Responsive	Responsive	
2	Continuous Passive Motion (Computerized) M/s. Bharat Medical Systems Make : Own	Non- responsive	Responsive	 The quoted model does not meet the following technical specification : 1) The limb support is not padded adequately and flimsy. 2) Ankle support not provided. 3) The length adjustment is not satisfactory. 4) Remote does not have angle adjustment function.
3	Height Adjusting Tilting Table M/s. Bharat Medical Systems Make : Own	Responsive	Responsive	