

RESULT OF EVALUATION AND FINALLY SHORTLISTED BIDDERS.

Tender No. 569/ELISA/RC/TNMSC/ENGG/2020, Dt.27.02.2020

TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND INSTALLATION OF ELISA READER & ELISA WASHER

Equipment: Elisa Reader

Sl. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
1	M/s. Shan biotech & Diagnostics	Robonik/ Readwell Touch	Responsive	Responsive	-----
2	M/s. Transasia Bio Medicals Ltd	Erba/Lisa Scan EM (4F)	Non Responsive	Responsive	Technical: As per specialist's report the model is not meeting the following points of the technical specification during the demonstration: 1. Optical System: 8seconds maximum measurement time. 2. Inputs: Touch and Membrane keyboard.

Equipment: Elisa Washer

Sl. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
1	M/s. Shan biotech & Diagnostics	Robonik/ Washwell Plate	Responsive	Responsive	-----
2	M/s. Transasia Bio Medicals Ltd	Erba/Lisa wash	Responsive	Responsive	-----