

**EVALUATION REPORT**

**Tender for fixing rate contract for supply and installation of  
Blood Bank Refrigerator, Blood Collection Monitor and Donor  
Blood Collection Couch to various Government Institutions**

**Tender No. 564/BBK/RC/TNMSCENGG/2020, Dt. 10.02.2020.**

-----

**a) Blood bank Refrigerator -20 Nos**

<b>Sc h. No</b>	<b>Name of the Equipment and Bidder</b>	<b>Demo report</b>	<b>Reason for rejection</b>
1	<b>M/s. CryoScientific Systems Pvt. Ltd.,</b> Model : BBR-160 Make : Cryo Scientific	Non-responsive	<p>The quoted model does not meet the following specification :</p> <ul style="list-style-type: none"> <li>a) Temperature- Some of the compartment did not achieve the desired temperature. Hold over time is less than 15 minutes.</li> <li>b) Controlling Resolution-0.5°C</li> <li>c) Chart Range - 0.40°C</li> <li>d) Display Resolution -0.5°C.</li> <li>e) Cabinet Material - External - 1.0mm sheet, powder-coated, Internal - 0.8mm SS Sheet.</li> <li>f) Voltage Stabilizer - A voltage stabilizer of suitable capacity is built-in.</li> <li>g) Ext. Dimensions (w x d x h) cm -60 x 62 x 180 approx.</li> <li>h) Int. Dimensions (w x d x h) cm -52 x 48.5 x 138 approx.</li> </ul>
2	<b>M/s. Asean Scientific express</b> Model : AS-120BBR Make : Asean scientific	Non-responsive	<p>The quoted model does not meet the following specification :</p> <ul style="list-style-type: none"> <li>a) Temperature- Some of the compartment did not achieve the desired temperature. Hold over time is less than 10 minutes.</li> <li>b) Controlling Resolution-0.5°C</li> <li>c) Chart Range - 0.40°C</li> <li>d) Display Resolution -0.5°C.</li> <li>e) Cabinet Material - External - 1.0mm sheet, powder-coated, Internal - 0.8mm SS Sheet.</li> <li>f) Voltage Stabilizer - A voltage stabilizer of suitable capacity is built-in.</li> <li>g) Ext. Dimensions (w x d x h) cm -60 x 62 x 180 approx.</li> <li>h) Int. Dimensions (w x d x h) cm -52 x 48.5 x 138 approx.</li> </ul>
3	<b>M/s. Alan Medical &amp; Laboratory Products</b> Model : BR 120 Make : REMI	Non-responsive	<p>The quoted model does not meet the following specification :</p> <ul style="list-style-type: none"> <li>a) Temperature- Some of the</li> </ul>

Sc h. No	Name of the Equipment and Bidder	Demo report	Reason for rejection
			<p>compartment did not achieve the desired temperature. Hold over time is less than 20 minutes.</p> <p>b) Voltage Stabilizer - A voltage stabilizer of suitable capacity is built-in.</p> <p>c) Door-3 pane heated Glass</p> <p>d) No. of Trays -5</p> <p>e) Tray capacity 24 (450 ml bags)</p> <p>f) Ext. Dimensions (w x d x h) cm -60 x 62 x 180 approx.</p> <p>Int. Dimensions (w x d x h) cm -52 x 48.5 x 138 approx.</p>
4	<b>M/s. Bioimplants India (P) Ltd.,</b> Make : Imperial Biotech	Non-responsive	Not demonstrated
5	<b>M/s. Solokrafts Industries</b> Model : EBR 300 Make : Elektrocraft	Non-responsive	Not demonstrated

#### b) Blood Collection Monitor – 20 Nos

Sc h. No	Name of the Equipment and Bidder	Demo report	Reason for rejection
1	<b>M/s. Alan Medical &amp; Laboratory Products</b> Model : BCM 10U Make : REMI	Non-responsive	<p>The quoted model does not meet the following specification :</p> <p>Oscillation details – Oscillation is below 10 rpm instead of 12 <math>\pm</math>2 rpm, motor driven oscillation.</p> <p>Not having IPXI certificate.</p> <p>Dimensions – 250 X 340 X 155 mm approx- Not matching</p>
2	<b>M/s. Bioimplants India (P) Ltd.,</b> Model : IBCM-03 Make : Imperial Biotech	Non-responsive	<p>The quoted model does not meet the following specification :</p> <p>Oscillation details – Oscillation is below 8 rpm instead of 12 <math>\pm</math>2 rpm, motor driven oscillation.</p> <p>Not having IPXI certificate.</p> <p>Dimensions – 250 X 340 X 155 mm approx- Not matching</p>
3	<b>M/s. Solokrafts Industries</b> Model : EDC 02 Make : Elektrocraft	Non-responsive	Not demonstrated

**c) Donor Blood Collection Couch – 20 Nos**

<b>Sc h. No</b>	<b>Name of the Equipment and Bidder</b>	<b>Demo report</b>	<b>Reason for rejection</b>
1	<b>M/s. Alan Medical &amp; Laboratory Products</b> Model : Desire 2X Make : REMI	Non-responsive	The quoted model does not meet the following technical specification :  a) Interface. b) Continuous operation. c) IPXO certificate - Not available. d) Upholstery – Soft upholstery of 2.5 inch thickness. e) BCM tray is very small. f) Length of fixed seat – 100 cm approx. g) Length of swivel seat– 90 cm approx. h) Width -67 cm approx i) Length of arm rest -60cm approx j) Width of arm rest -15cm approx k) Height of seat -51 cm approx
2	<b>M/s. Bioimplants India (P) Ltd.,</b> Make : Imperial Biotech	Non-responsive	The quoted model does not meet the following technical specification :  a) Interface. b) Classification- Protection against electrical shock class I c) Continuous operation. d) IPXO certificate - Not available. e) Intimation of motor movement – By feather touch key pad f) Width -67 cm approx g) Length of arm rest -60cm approx h) Width of arm rest -15cm approx Height of seat -51 cm approx
3	<b>M/s. Solokrafts Industries</b> Model : EDC 02 Make : Elektrocraft	Non-responsive	The quoted model does not meet the following technical specification :  a) Interface. b) Classification- Protection against electrical shock class I c) Continuous operation. d) IPXO certificate - Not available. e) No electronically adjustable motors. f) Intimation of motor movement – By feather touch key pad. l) Upholstery – Soft upholstery of 2.5 inch thickness. g) BCM tray is very small. m) Length of fixed seat – 100 cm approx. n) Length of swivel seat– 90 cm approx. h) Width -67 cm approx i) Length of arm rest -60cm approx j) Width of arm rest -15cm approx Height of seat -51 cm approx

**"THE TENDER IS CANCELLED"**