

**INTERNATIONAL COMPETITIVE BIDDING (ICB) FOR SUPPLY AND
INSTALLATION OF DIGITAL MAMMOGRAPHY SYSTEM FOR MEDICAL
COLLEGE HOSPITALS**

**TENDER NO. PKG2/ICB/TNUHP/ JICA/C2/MAM/TNMSC/ENGG/ 2020
dated 04.02.2020**

RESULT OF EVALUATION

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection / Remarks
		Technical Responsive ness	Commercial Responsiveness	
Digital Mammography System				
1.	M/s Trivitron Healthcare Pvt. Ltd., Make: Hologic Model: Selenia Dimensions 3D Mammo System	Responsive	Responsive	-----
2.	M/s Panacea Medical Technologies Private Ltd., Make: Panacea Model: Lilac	Non - Responsive	Responsive	<u>Technical terms:</u> As per specialist remarks the quoted model has deviations for the following technical specifications: i) control buttons for adjustment of height and angles shall be operable from gantry as well as from foot paddles. ii) Grid ration shall be 5:1 or better with atleast 30 line pairs per cm. iii) Motorized installation and removal/auto retract of grid/breast support assembly system shall be available for geometric magnification. iv) No upgradable software. <u>Commercial terms:</u> Documents meeting minimum qualification criteria not furnished as per tender requirement. i) Production capacity for Last 5 years rom 2014-15 to 2018-19. ii) Annual turnover does not meet minimum requirement. iii) Performance satisfactory reports furnished does not meet minimum requirement.
3.	M/s. Allengers Medical Systems Ltd. Make: Allengers Model: MAM Venus +	Non - Responsive	Responsive	<u>Technical terms:</u> As per specialist remarks the quoted model has deviations for the following technical specifications: i) Displayed parameters – Density selection ii) Paddle with shifting functionality that can be positioned left, right and middle of the detector such that the

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection / Remarks
		Technical Responsiveness	Commercial Responsiveness	
				shift of the X-Ray field is interlocked with shift of compression plate. iii) Image matrix in pixel shall be atleast 2800x3500 pixels or better. iv) Histogram display.
4.	M/s. Fujifilm India Private Limited Make: Fujifilm Model: AMULET Innovality	Responsive	Responsive	-----
5.	M/s Boston IVY Healthcare Solutions Pvt. Ltd., Make: United Imaging Model: U Mammo 890i	Non - Responsive	Non - Responsive	<u>Technical terms:</u> As per specialist remarks the quoted model has deviations for the following technical specifications: i) control buttons for adjustment of height and angles shall be operable from gantry as well as from foot paddles. ii) Motorized installation and removal/auto retract of grid/breast support assembly system shall be available for geometric magnification. iii) 1.8 or 1.6 magnification attachment with field magnification and spot magnification paddles – minimum of 2 paddles. iv) No upgradable software. <u>Commercial terms:</u> Documents meeting minimum qualification criteria not furnished as per tender requirement. ii) Annual Turnover for Bidder does not meet minimum requirement and the following forms are furnished: i) Form CON (Manufacturer) ii) Form FIN-1 iii) Form FIN-2 (Manufacturer) iv) Form FIR – 1 v) Form EXP-2 and relevant documents vi) Form EXP-3

Sd/-
General Manager (E)