

Date: 13.01.2021

IFB NO. PKG5/ICB/TNUHP/JICA/C2/RIG/TNMSC/ENGG/2020
dt 04.02.2020, for 11 nos. of FESS Complete Set with Microdebrider
(Lot/Contract no.1) and 8 nos. of Laparoscope (High definition)
for Medical College Hospitals under JICA

RESULT OF EVALUATION

Sl. No	Bidders Quoted	Equipment Details	Responsiveness		Reason for rejection / Remarks
			Technical Responsiveness	Commercial Responsiveness	
FESS Complete Set with Microdebrider (Lot/Contract no.1)					
1.	M/s. Vishal Surgical Equipment Co. Pvt. Ltd., Chennai	<p>Make: Karlstorz</p> <p>Model:</p> <p><u>A) FESS Complete Set:</u></p> <p>a) <u>High Definition Camera:</u></p> <p>(i) <u>3-chip Full HD Camera:</u> Image I Connect H3-Z Spies</p> <p>(ii) <u>Connect Module:</u> Image I Connect</p> <p>(iii) <u>Link Module:</u> Image I H3-Link</p> <p>b) High Definition Medical Grade Monitor: 27" Full HD Monitor</p> <p>c) Xenon Light Source with Fiber optic cable: Cold Light Fountain Xenon Nova 300 (300W)</p> <p>d) Endoscopic Trolley: Indian make</p> <p><u>B) Microdebrider:</u></p> <p>a) Microdebrider: Undrive SIII</p> <p>b) Multifunctional Foot Switch: Two pedal footswitch</p>	Responsive	Responsive	---

Sl. No	Bidders Quoted	Equipment Details	Responsiveness		Reason for rejection / Remarks
			Technical Responsiveness	Commercial Responsiveness	
		c) Micro Debrider Autoclavable: Drill Cut-X II Shaver Handpiece			
<u>Laparoscope (High definition) (Lot/Contract no.2)</u>					
1.	M/s. Vishal Surgical Equipment Co. Pvt. Ltd., Chennai	Make: Karl Storz Model: HOPKINS a) <u>Telescope:</u> (i) <u>0 deg:</u> 26003ACA ii) <u>30 deg (length 29cm):</u> 26046BCA iii) <u>10 mm 30 deg length 31cm):</u> 26003BCA b) <u>Reusable hand instruments:</u> Storz c) <u>Electronic CO2 insufflator:</u> Endoflator 40 (40L) d) <u>3-chip Full HD Camera:</u> Image I Connect H3-Z Spies Processor: Image I Connect and Image I H3-Link e) <u>HD Monitor with stand:</u> 27" FULL HD Monitor f) <u>Light Source:</u> Cold Light Fountain D-Light P SCB (300W Xenon) g) <u>Trolley:</u> Indian Make	Responsive	Responsive	---
2.	M/s. Indian Orthocare Surgicals, Chennai	Make: Stryker Model: Precision Ideal Eyes (IE) a) <u>Telescope:</u> i) <u>0 deg:</u> 0502-103-010 ii) <u>30 deg (length 29cm):</u> 0502-503-030 iii) <u>10 mm 30 deg length 31cm):</u> 0502-937-030	Responsive	Responsive	-----

Sl. No	Bidders Quoted	Equipment Details	Responsiveness		Reason for rejection / Remarks
			Technical Responsiveness	Commercial Responsiveness	
		b) <u>Reusable hand instruments:</u> Kare Lap instruments c) <u>Electronic CO2 insufflator:</u> Pneumosure (45L) d) <u>HD 3 chip Camera:</u> 1588 Advanced Imaging Modality (AIM) and Processor- 18mm Coupler and HD Zoom Coupler e) <u>HD Monitor with stand:</u> Vision pro 26" LED f) <u>Light Source:</u> L10 LED (400W) g) <u>Trolley:</u> IOS Kare Laparoscope Trolley			
3.	M/s Olympus Medical Systems India Pvt. Ltd., Chennai	Make: Olympus, Japan Model: VISERA ELITE II a) <u>Telescope:</u> i) <u>0 deg:</u> IR TELESCOPE WAIR100A ii) <u>30 deg (length 29cm):</u> IR TELESCOPE WAIR530A, 5.4mm iii) <u>10 mm 30 deg length 31cm):</u> IR TELESCOPE WAIR130A, 10mm b) <u>Reusable hand instruments:</u> HiQ+ Series c) <u>Electronic CO2 insufflator:</u> UHI-4	Non-Responsive	Responsive	<u>Reasons for Rejection:</u> <u>Technical Points:</u> <u>(ii) Re-usable Hand Instruments:</u> g) Reddick olsen dissecting and grasping forceps with 360 degree rotation, it shall have a connector pin for unipolar coagulation. Size 5mm and length 36cm . The complete instruments shall

Sl. No	Bidders Quoted	Equipment Details	Responsiveness		Reason for rejection / Remarks
			Technical Responsiveness	Commercial Responsiveness	
		<p>d) <u>HD 3 chip</u> <u>Camera:</u>CH-S200-XZ-EB</p> <p>Processor: OTV-S200</p> <p>e)<u>HD Monitor with</u> <u>stand:</u>LMD-2735 SONY HD</p> <p>f) <u>Light Source:</u>CLV-S200 IR XENON</p> <p>g) <u>Trolley:</u> A10000000410, Indian Make</p>			<p>consist of: A plastic handle without ratchet, an insulated outer tube and an insert - 1No.: 33 cm only available.</p> <p>h)Toothed 2*4 teeth Grasping forceps with 360 degree rotation, it shall have a connector pin for unipolar coagulation. Size 5mm and length 36cm. The complete instruments shall consist of: A plastic handle with ratchet, an insulated outer tube and an insert - 1No.:33 cm only available</p> <p>i)Metzenbaum curved scissors 360 degree rotation and single action jaws, it shall have a connector pin</p>

Sl. No	Bidders Quoted	Equipment Details	Responsiveness		Reason for rejection / Remarks
			Technical Responsiveness	Commercial Responsiveness	
					<p>for unipolar coagulation. Size 5mm and length 36cm. The complete instruments shall consist of: A plastic handle without ratchet, an insulated outer tube and an insert - 1No.:33 cm only available</p> <p>k)Dissecting cannula, L - shaped size 3mm working length 36cm - 2 No.:33 cm only available</p> <p>l)Dissecting spatula, blunt size 5mm working length 36cm - 2 No.:33 cm only available</p> <p>m)Bipolar forceps, size 5mm length 36cm with connector pin for bipolar coagulation. The</p>

Sl. No	Bidders Quoted	Equipment Details	Responsiveness		Reason for rejection / Remarks
			Technical Responsiveness	Commercial Responsiveness	
					complete set shall consist of a outer tube bipolar forceps, a spring handle with luer - lock connector for easy cleaning and an insert with flat jaws- 1No.: 33 cm only available

Sd/-
General Manager (E)