

Date: 13.01.2021

IFB NO. PKG8/ICB/TNUHP/JICA/C2/HEM/TNMSC/ENGG/2020

dt 04.02.2020, for 16 nos. of CRRT Complete set

(Lot/Contract no.1)for Medical College Hospitals under JICA

RESULT OF EVALUATION

Sl. No	Bidders Quoted	Equipment Details	Responsiveness		Reason for rejection / Remarks
			Technical Responsiveness	Commercial Responsiveness	
CRRT Complete Set (Lot/Contract no.1)					
1.	M/s Baxter (India) Pvt. Ltd., Gurgaon	Make: <i>Baxter, Italy</i> Model: <i>Prisma Flex 8.1</i>	Responsive	Responsive	---
2.	M/s Fresenius Medical Care India Pvt. Ltd., Chennai	M/s Make: <i>Fresenius, Germany</i> Model: <i>Multi filtrate</i>	Non-Responsive	Non-Responsive	Reasons for Rejection: Technical Points: vi) Other Features: c) "Option for simultaneous delivery of Pre and post filter replacement solution in CVVH (Continuous Venovenous hemofiltration) and CVVHFD (Continuous Venovenous hemodialysis). " (vii) Consumables-30 kits. Commercial Points: Manufacturer: a) Form ACK: Not Furnished b) Form FIN- :Not Furnished c) Form FIN- 2:Not Furnished

Sl. No	Bidders Quoted	Equipment Details	Responsiveness		Reason for rejection / Remarks
			Technical Responsiveness	Commercial Responsiveness	
					d) Form FIR-1:Not Furnished e) Form FIR-2:Not Furnished f) Form EXP-1:Not Furnished g) Form EXP-2:Not Furnished h) Form EXP-3:Not Furnished

Sd/-
General Manager (E)