

Date: 25.01.2021

**IFB NO. PKG5/ICB/TNUHP/JICA/C2/RIG/TNMSC/ENGG/2020  
dt 04.02.2020, for 11 nos. of FESS Complete Set with Microdebrider  
(Lot/Contract no.1) and 8 nos. of Laparoscope (High definition)  
for Medical College Hospitals under JICA  
RESULT OF RE-EVALUATION**

Sl. No	Bidders Quoted	Equipment Details	Responsiveness		Reason for rejection / Remarks
			Technical Responsiveness	Commercial Responsiveness	
FESS Complete Set with Microdebrider (Lot/Contract no.1)					
1.	M/s. Vishal Surgical Equipment Co. Pvt. Ltd., Chennai	<p><b>Make:</b> Karlstorz</p> <p><b>Model:</b></p> <p><b><u>A) FESS Complete Set:</u></b></p> <p><b>a) <u>High Definition Camera:</u></b></p> <p><b>(i) <u>3-chip Full HD Camera:</u></b> Image I Connect H3-Z Spies</p> <p><b>(ii) <u>Connect Module:</u></b> Image I Connect</p> <p><b>(iii) <u>Link Module:</u></b> Image I H3-Link</p> <p><b>b) High Definition Medical Grade Monitor:</b> 27" Full HD Monitor</p> <p><b>c) Xenon Light Source with Fiber optic cable:</b> Cold Light Fountain Xenon Nova 300 (300W)</p> <p><b>d) Endoscopic Trolley:</b> Indian make</p> <p><b><u>B) Microdebrider:</u></b></p> <p>a) <b>Microdebrider:</b> Undrive SIII</p> <p>b) <b>Multifunctional Foot Switch:</b> Two pedal footswitch</p> <p>c) <b>Micro Debrider Autoclavable:</b> Drill Cut-X II Shaver Handpiece</p>	Responsive	Responsive	---

Sl. No	Bidders Quoted	Equipment Details	Responsiveness		Reason for rejection / Remarks
			Technical Responsiveness	Commercial Responsiveness	
Laparoscope (High definition) (Lot/Contract no.2)					
1.	M/s. Vishal Surgical Equipment Co. Pvt. Ltd., Chennai	<b>Make:</b> Karl Storz <b>Model:</b> HOPKINS <b>a) <u>Telescope:</u></b> <b>(i) <u>0 deg:</u></b> 26003ACA <b>ii) <u>30 deg (length 29cm):</u></b> 26046BCA <b>iii) <u>10 mm 30 deg length 31cm):</u></b> 26003BCA <b>b) <u>Reusable hand instruments:</u></b> Storz <b>c)<u>Electronic CO2 insufflator:</u></b> Endoflator 40 (40L) <b>d) <u>3-chip Full HD Camera:</u></b> Image I Connect H3-Z Spies <b>Processor:</b> Image I Connect and Image I H3-Link <b>e)<u>HD Monitor with stand:</u></b> 27" FULL HD Monitor <b>f) <u>Light Source:</u></b> Cold Light Fountain D-Light P SCB (300W Xenon) <b>g) <u>Trolley:</u></b> Indian Make	Responsive	Responsive	---
2.	M/s. Indian Orthocare Surgicals, Chennai	<b>Make:</b> Stryker <b>Model:</b> Precision Ideal Eyes (IE) <b>a) <u>Telescope:</u></b> <b>i) <u>0 deg:</u></b> 0502-103-010 <b>ii) <u>30 deg (length 29cm):</u></b> 0502-503-030 <b>iii) <u>10 mm 30 deg length 31cm):</u></b> 0502-937-030 <b>b) <u>Reusable hand instruments:</u></b> Kare Lap instruments	Responsive	Responsive	-----

Sl. No	Bidders Quoted	Equipment Details	Responsiveness		Reason for rejection / Remarks
			Technical Responsiveness	Commercial Responsiveness	
		<b>c) <u>Electronic CO2 insufflator:</u></b> Pneumasure (45L) <b>d) <u>HD 3 chip Camera:</u></b> 1588 Advanced Imaging Modality (AIM) and <b>Processor-</b> 18mm Coupler and HD Zoom Coupler <b>e) <u>HD Monitor with stand:</u></b> Vision pro 26" LED <b>f) <u>Light Source:</u></b> L10 LED (400W) <b>g) <u>Trolley:</u></b> IOS Kare Laparoscope Trolley			
3.	M/s Olympus Medical Systems India Pvt. Ltd., Chennai	<b>Make:</b> Olympus, Japan <b>Model:</b> VISERA ELITE II <b>a) <u>Telescope:</u></b> <b>i) <u>0 deg: IR TELESCOPE WAIR100A</u></b> <b>ii) <u>30 deg (length 29cm): IR TELESCOPE WAIR530A, 5.4mm</u></b> <b>iii) <u>10 mm 30 deg length 31cm): IR TELESCOPE WAIR130A, 10mm</u></b> <b>b) <u>Reusable hand instruments:</u></b> HiQ+ Series <b>c) <u>Electronic CO2 insufflator:</u></b> UHI-4 <b>d) <u>HD 3 chip Camera:</u></b> CH-S200-XZ-EB <b>Processor:</b> OTV-S200 <b>e) <u>HD Monitor with stand:</u></b> LMD-2735 SONY HD <b>f) <u>Light Source:</u></b> CLV-S200 IR XENON <b>g) <u>Trolley:</u></b> A10000000410, Indian Make	Responsive	Responsive	<b><u>Specialist's remarks:</u></b> 1. Laparoscopic instruments (Make: HIQ Series) are outsourced and the shaft length of hand instruments are 330mm and along with jaw insert it comes to 360mm, which meet the tender specifications.

**Sd/-  
General Manager (E)**