

03.02.2021

Evaluation Report

Tender no. 534/SCOPE/DMS/TNMSC/ENGG/2019, Dt. 11.10.2019

Tender for fixing Rate Contract for Supply and Installation of Basic Diagnostic Arthroscopy set to various Govt. Institutions

Sch. No. (1) : Basic Diagnostic Arthroscopy Set – 1 no.

Sch . No	Name of the Bidder Make/Model	Responsiveness		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	<p>M/s. Avana Medical Devices Pvt. Ltd. Option I Model : 4K Camera System –Arthrex Scopes, Shaver Pump & all instruments – Arthrex, GmbH Electrical Drill / Saw – Indigenous</p> <p>Option II Model : 4K Camera System –Arthrex Scopes, Shaver Pump & all instruments – Arthrex (Without Electrical Drill System)</p> <p>Option III Model : HD Camera System –Oumen Medical (M/s. Jiangsu Ouman Electronic Equipment co., Ltd.) Scopes, Shaver Pump & all instruments – Arthrex Electrical Drill / Saw – Indigenous</p> <p>Option IV Model : HD Camera System –Richard Wolf (M/s. RIWOspine) Scopes, Shaver Pump & all instruments – Arthrex Electrical Drill / Saw – Indigenous</p>	<p>Option I & Option II Non Responsive</p> <p>Option III & Option IV Responsive</p>	<p>Option I & Option II Non Responsive</p> <p>Option III & Option IV Responsive</p>	<p>Option I & Option II As per Corrigendum dt. 01.10.2020, the technical specification of 4K Camera System has been deleted, the model quoted for 4K Camera System has not been recommended</p>
2	<p>M/s. Indian Orthocare Surgicals Model : Arthroscopy System ; Stryker, USA Endoscopy Trolley : KARE IOS make 1488 HD Camera System L9000 LED Light Source; Precision LED Light source CORE Arthroscopic Shaver System VisionPro 26" LED Display Make : Stryker. USA</p>	Responsive	Responsive	-----

Sch . No	Name of the Bidder Make/Model	Responsiveness		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
3	<p>M/s. Vishal Surgical Equipment Co. Pvt. Ltd.</p> <p>Option I Model : Hopkins Telescope UniDrive S III Arthro SCB; Drillcut-X Arthro Shaver Handpiece; Image I Connect module with integrated Karl Storz-SCB; Image 1 H3 Link; 27" Full HD monitor; Karl Storz Xenon Nova 300 Light source; Indian make trolley Make : Karl Storz, Germany</p> <p>Option II Model : Hopkins Telescope; Image I Connect module with integrated Karl Storz-SCB; Image 1 H3 Link; 27" Full HD monitor; Karl Storz Xenon Nova 300 Light source; Indian make trolley (without UniDrive S III Arthro SCB; Drillcut-X Arthro Shaver Handpiece for Sl. No. 10) Make : Karl Storz, Germany</p>	<p>Option I & Option II Responsive</p>	<p>Option I & Option II Responsive</p>	<p>-----</p>