

**INTERNATIONAL COMPETITIVE BIDDING (ICB) FOR SUPPLY AND  
INSTALLATION OF RADIOLOGY MEDICAL EQUIPMENT FOR MEDICAL  
COLLEGE HOSPITALS AND DISTRICT HEADQUARTERS HOSPITAL**

**TENDER NO. PKG1/ICB/TNUHP/ JICA/C2/RAD/TNMSC/ENGG/ 2020  
dated 04.02.2020**

**RESULT OF EVALUATION**

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection / Remarks
		Technical Responsiveness	Commercial Responsiveness	
Lot/ Contract 1: Digital X-Ray Machine				
1.	M/s. Carestream Health India Pvt. Ltd.	Responsive	Substantially Responsive	-----
2.	M/s. Prognosys Medical Systems Pvt. Ltd.	Responsive	Responsive	-----
3.	M/s. Allengers Medical Systems Ltd.	Non - Responsive	Responsive	<u>Technical terms:</u> As per specialist remarks the quoted model has deviations for the following technical specifications: i) Digital display of all radiographic parameters ii) Minor KVp control iii) mobile four way positioning of the patient.
Lot/ Contract 2: Digital Radiography with fluoroscopy				
1.	M/s. Shimadzu Medical India Pvt. Ltd.	Responsive	Responsive	-----
2.	M/s. Allengers Medical Systems Ltd.	Non - Responsive	Responsive	<u>Technical terms:</u> As per specialist remarks the quoted model has deviations for the following technical specifications: i) Remote controlled RF table with tilting and elevation. ii) LED readout for exposure parameter display

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection / Remarks
		Technical Responsiveness	Commercial Responsiveness	
				iii) 19" or more monochrome LCD Medical grade iv) light weight less than 8 kg.
<b>Lot/ Contract 3: Mobile X-Ray Machine</b>				
1.	M/s. Allengers Medical Systems Ltd.	Responsive	Responsive	-----
<b>Lot/ Contract 4: Computed Radiography Machine</b>				
1.	M/s. Carestream Health India Pvt. Ltd.	Responsive	Substantially Responsive	-----
2.	M/s. Fujifilm India Private Limited	Responsive	Responsive	-----
<b>Lot/ Contract 5: C-ARM Machine</b>				
1.	M/s. Shimadzu Medical India Pvt. Ltd.	Responsive	Substantially Responsive	-----
2.	M/s. Allengers Medical Systems Ltd.	Non - Responsive	Responsive	<u>Technical terms:</u> As per specialist remarks the quoted model has deviations for the following technical specifications: i) Fully balanced c-ARM with electro-magnetic lock mechanism for individual locking of all movements. ii) X-Ray image shall be available on the control panel in LCD display. iii) All C-ARM movements to be locked/ unlocked using single touch operation iv) Video Printer.
3.	M/s. Wipro GE Healthcare Private Limited	Non - Responsive	Responsive	<u>Technical terms:</u> As per specialist remarks the quoted model has deviations for

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection / Remarks
		Technical Responsiveness	Commercial Responsiveness	
				<p>the following technical specifications:</p> <p>i) Fully balanced c-ARM with electro-magnetic lock mechanism for individual locking of all movements.</p> <p>ii) Radiographic mAs value: minimum 150mAs.</p> <p>iii) Collimation position memory</p> <p>iv) All C-ARM movements to be locked/ unlocked using single touch operation</p> <p>v) C-ARM handle with lock in the image intensifier to be provided.</p>
4.	M/s. Trivitron Healthcare Pvt. Ltd.	Responsive	Responsive	-----

**Sd/-**

**General Manager (E)**