

16.02.2021

## Evaluation Report

**Tender no. 580/SCOPE/DME-TRY/TNMSC/ENGG/2020, Dt. 17.06.2020**

Tender for Supply and Installation of Fiber Optic Bronchoscope, Rigid  
Bronchoscope to Dept. of Tuberculosis and Chest Disease and  
Spectrophotometer to Dept. of BioChemistry to KAP Viswanathan Govt.  
Medical College Hospital, Trichy

### Sch. No. (2) : Rigid Bronchoscope – 1 no.

Sch . No	Name of the Bidder Make/Model	Responsiveness		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Olympus Medical Systems India Pvt. Ltd. Model : LTF-160 (Evis Exera Pleuravideoscope), LMD 2435 (SONY) HD Monitor, FB-20C-1, NA-411D-1521, BC- 202D-3010, MAJ-1058(E), MAJ-1430, Trolley Indian Make Software accessories Make : Olympus, Japan	Responsive	Responsive	-----

### Sch. No. (3) : Spectrophotometer – 1 no.

Sch . No	Name of the Bidder Make/Model	Responsiveness		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Alan Medical and Laboratory Products Model : LMSP- UV1900 Make : Labman, India	Non Responsive	Non Responsive	<b>Technical Evaluation:</b> The quoted model is not meeting the below technical specifications: a. Instead of light Source : Xenon Flash Lamp, the quoted model has D2 and Tungsten Halogen lamp. b. Sample Holder : Should have reference and sample cuvette positions, should have provisions to dismantle and attach multiple capacity holders – Not available.

Sch . No	Name of the Bidder Make/Model	Responsiveness		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
				<p>c. Cuvettes : glass 10 Nos. and quartz 10 nos. of variable capacities – Not available.</p> <p>d. Cell holder : Automatic 6 cell turret – Not mentioned</p> <p>e. Interface with latest version of branded computer and Laser Printer – Laser printer not mentioned.</p> <p><b>Commercial Evaluation:</b> Did not furnish Tender terms and conditions (Validity, Delivery schedule, Payment, Warranty, Spare parts), Manufacturing capacity, PO, Installation report &amp; End user certificate for manufacturer for meeting the qualification criteria on past supply, hence not recommended.</p>
2	M/s. Inexus Biotech Pvt. Ltd. Model : UH-5300 Make : Hitachi, Japan	Responsive	Responsive	-----

**Sd/-  
General Manager (E)**