

**Tender for supply and Installation of Echo 2D Color Doppler to Dept. of
Cardiology at Rajiv Gandhi Govt. General Hosptial, Chennai
Tender No.572/USG/RGGGH/TNMSC/ENGG/2020, Dt.17.06.2020**

RESULTS OF EVALUATION

Echo 2D Color Doppler:-

S. No	Name of the bidder	Technical Evaluation report	Commercial evaluation report	Reason for Rejection
1	M/s. Wipro GE Healthcare Pvt Ltd Model: GE VIVID 60 Make: Own	Non – Responsive	Responsive	In technical points: 1. No information on number of channels in page no.5 of the P.D Sheet, technical deviation (Point No. 2) 2. System maximum depth is 36 cm only (Point No.14) 3. System frame rate 1500 frames/second (Point No. 15) 4. 6Tc-RS TEE probe, doesn't mention number of elements, as per spec it should be 2500 elements (Point No. 19 (c))
2	M/s. Sri Sai Ekdant Surgicals Model: Affiniti 50 Make: Philips	Responsive	Responsive	-----
3	M/s. Esaote India (NS) Ltd Option 1 Model: MyLab 9 Make: Own	Non – Responsive	Responsive	In technical points: 1. Touch panel measures 10.1” (Point No.4) 2. ST 2612 TEE Adult, number of elements, as per sepc it should be 2500 elements – No details about no. of elements (Point No. 19 (c))
4	M/s. Esaote India (NS) Ltd Option 2 Model: MyLab X8 Make: Own	Non – Responsive	Responsive	In technical points: 1. ST 2612 TEE Adult, number of elements, as per sepc it should be 2500 elements – No details about no. of elements (Point No. 19 (c))