

01.03.2021

**INTERNATIONAL COMPETITIVE BIDDING (ICB) FOR SUPPLY AND
INSTALLATION OF RADIOLOGY MEDICAL EQUIPMENT FOR MEDICAL
COLLEGE HOSPITALS AND DISTRICT HEADQUARTERS HOSPITAL**

**TENDER NO. PKG1/ICB/TNUHP/ JICA/C2/RAD/TNMSC/ENGG/ 2020
dated 04.02.2020**

RESULT OF EVALUATION

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection / Remarks
		Technical Responsive ness	Commercial Responsive ness	
Lot/ Contract 1: Digital X-Ray Machine				
1.	M/s. Carestream Health India Pvt. Ltd.	Responsive	Substantially Responsive	-----
2.	M/s. Prognosys Medical Systems Pvt. Ltd.	Responsive	Responsive	-----
3.	M/s. Allengers Medical Systems Ltd.	Substantially Responsive	Responsive	-----
Lot/ Contract 2: Digital Radiography with fluoroscopy				
1.	M/s. Shimadzu Medical India Pvt. Ltd.	Responsive	Responsive	-----
2.	M/s. Allengers Medical Systems Ltd.	Substantially Responsive	Responsive	-----
Lot/ Contract 3: Mobile X-Ray Machine				
1.	M/s. Allengers Medical Systems Ltd.	Responsive	Responsive	-----
Lot/ Contract 4: Computed Radiography Machine				
1.	M/s. Carestream Health India Pvt. Ltd.	Responsive	Substantially Responsive	-----
2.	M/s. Fujifilm India Private Limited	Responsive	Responsive	-----

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection /
		Technical Responsive ness	Commercial Responsive ness	Remarks
Lot/ Contract 5: C-ARM Machine				
1.	M/s. Shimadzu Medical India Pvt. Ltd.	Responsive	Substantially Responsive	-----
2.	M/s. Allengers Medical Systems Ltd.	Non - Responsive	Responsive	<u>Technical terms:</u> As per specialist remarks the quoted model has deviations for the following technical specifications: i) Fully balanced c-ARM with electro-magnetic lock mechanism for individual locking of all movements. ii) X-Ray image shall be available on the control panel in LCD display. iii) All C-ARM movements to be locked/ unlocked using single touch operation iv) Video Printer.
3.	M/s. Wipro GE Healthcare Private Limited	Non - Responsive	Responsive	<u>Technical terms:</u> As per specialist remarks the quoted model has deviations for the following technical specifications: i) Fully balanced c-ARM with electro-magnetic lock mechanism for individual locking of all movements. ii) Radiographic mAs value: minimum 150mAs.

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection / Remarks
		Technical Responsive ness	Commercial Responsive ness	
				iii) Collimation position memory iv) All C-ARM movements to be locked/ unlocked using single touch operation v) C-ARM handle with lock in the image intensifier to be provided.
4.	M/s. Trivitron Healthcare Pvt. Ltd.	Responsive	Responsive	-----

**Sd/-
General Manger (E)**