

21.03.2022

EVALUATION RESULT

Tender for fixing rate contract for supply and installation of fully automated Haematology analyzer (3 & 5 part) to various Government Medical Institutions

Tender ref. 549/CELL/RC/TNMSC/ENGG/2020, Dt. 09.01.2020.

Sch. No. 1:- Fully Automated Hematology Analyzer (3 part) / Cell Counter (Type-I with Neutrophils)

Name of the bidder with model & make	Responsiveness		Reason for rejection if any
	Technical Evaluation	Commercial Evaluation	
M/s. Sysmex India Pvt. Ltd., Mumbai Model: XP-300 Make Sysmex	Responsive	Responsive	Nil
M/s. Trivitron Healthcare Pvt. Ltd., Chennai Model: Cellenium 21 Make: Trivitron	Non - responsive	Responsive	a) Neutrophils cannot be detected as measured parameter. b) No built in software to connect with the existing printer provided by the institute.
M/s. Rapid Diagnostics Pvt. Ltd., Delhi Model: BC-21s Make: Mindray	Non-responsive	Responsive	a) No LCD display only TGT display (lesser resolution). b) HCT is a calculated parameter and not a measured parameter. c) Calibration factor not shown for HCT in type 1. d) Neutrophils cannot be detected as measured parameter. e) No built in software to connect with the existing printer provided by the institute.

Sch. No. 2:- Fully Automated Hematology Analyzer (3 part) / Cell Counter (Type-II with Granulocytes)

Name of the bidder with model & make	Responsiveness		Reason for rejection if any
	Technical Evaluation	Commercial Evaluation	
M/s. Siemens Healthcare Pvt. Ltd., Chennai Model: Advia 360 Make:- Siemens	Responsive	Responsive	NIL
M/s. Transasia Bio-Medicals Ltd., Mumbai Model: H 360 Make:- Erba	Responsive	Responsive	NIL
M/s. Trivitron Healthcare Pvt. Ltd., Chennai Model: Cellenium 21 Make: Trivitron	Non - responsive	Responsive	No built in software to connect with the existing printer provided by the institute.
M/s. Rapid Diagnostics Pvt. Ltd., Delhi	Non - responsive	Responsive	1) No LCD display only TGT display (lesser resolution).

Name of the bidder with model & make	Responsiveness		Reason for rejection if any
	Technical Evaluation	Commercial Evaluation	
Model: BC-21s Make: Mindray			2) Calibration factor not shown for HCT in type 1. 3) No built in software to connect with the existing printer provided by the institute.

Sch. No. 3:- Fully Automated Hematology Analyzer (5 part) / Cell Counter

Name of the bidder with model & make	Responsiveness		Reason for rejection if any
	Technical Evaluation	Commercial Evaluation	
M/s. Siemens Healthcare Pvt. Ltd., Chennai Model: Advia 560 & 2120i Make:- Siemens	Substantially - responsive	Responsive	Nil
M/s. Sysmex India Pvt. Ltd., Mumbai Model: XN-350 Make Sysmex	Responsive	Responsive	Nil
M/s. Deep Meditech Pvt. Ltd., Delhi Model: BC 6200 Make: Mindray	Non - responsive	Responsive	a) Histogram for WBC not available. b) Separate QC for reticulocyte against the requirement of single QC material.
M/s. Transasia Bio-Medicals Ltd., Mumbai Model: H 560 Make:- Erba	Non - responsive	Responsive	a) Not capable to process Reticulocyte, body fluid counts and IPF to measure immature platelet fraction.

**Sd/-
General Manager (E)**