

Tender no.617/VEN/NHM/TNMSC/ENGG/2020, Dt. 21.10.2020

Tender for supply and installation on Neonatal ventilator with High Frequency Oscillatory to various Govt. Institutions

RESULT OF EVALUATION AND FINALLY SHORTLISTED BIDDERS

Sch. No. (1):- Neonatal ventilator with High Frequency Oscillatory Ventilation - 3 nos.

Sc h. No	Name of the Bidder Make/Model	Responsiveness		Reason for rejection
		Technical evaluation report	Commerci al evaluation report	
1	M/s. Madras Surgicals & Chemicals Make : Draeger Model : BABYLOG VN600	Responsive	Responsive	-----
2.	M/s. Trans healthcare India Pvt. Ltd. Make : Vyaire Medical INC, USA Model : Fabian HFO	Not Responsive	Responsive	1. Screen size is varied from the given specification. 2. Compressor will be provide only on request. 3. Scalar Graphs & Loops are displayed in separate windows. 4. Mandatory minute ventilation not available. 5. Proportion pressure support not available. 6. Automatic tube compensation not available 7. Volume Guarantee for HNOV only facility for volume limiting available. 8. Help function not available