

## **EVALUATION RESULT**

**TENDER NO.719/MRI/RC/TNMSC/ENGG/2021, DT. 07.09.2021**

### **TENDER FOR FIXING RATE CONTRACT FOR 1.5 TESLA MRI SYSTEM TO VARIOUS MEDICAL INSTITUTIONS IN THE STATE OF TAMILNADU**

**Sch. 1) Option I-A (10KW/8 channel/30mt/m& 100 T/m/s):**

<b>Sl. No</b>	<b>Name of the firm</b>	<b>Technical evaluation report</b>	<b>Commercial evaluation report</b>	<b>Reason for rejection</b>
1	M/s. Siemens Healthcare Private Limited MAGNETOM Semptra 8 channel	Responsive	Responsive	-----
2.	M/s. Philips India Ltd. Prodiva 1.5T CS	Responsive	Responsive	-----

**Sch. 1) Option I-B (10KW/8 channel/33 mt/m& 120 T/m/s):**

<b>Sl. No</b>	<b>Name of the firm</b>	<b>Technical evaluation report</b>	<b>Commercial evaluation report</b>	<b>Reason for rejection</b>
1	M/s.Wipro GE Healthcare Pvt. Ltd GE Signa Creator – 8 channel	Responsive	Responsive	-----
2.	M/s. Philips India Ltd. Prodiva 1.5T CS	Responsive	Responsive	-----

**Sch. 2) Option II-A (10KW/16 channel/ 30mt/m& 100 T/m/s)**

<b>Sl. No</b>	<b>Name of the firm</b>	<b>Technical evaluation report</b>	<b>Commercial evaluation report</b>	<b>Reason for rejection</b>
1	M/s. Siemens Healthcare Private Limited MAGNETOM Semptra 16 channel	Responsive	Responsive	-----
2.	M/s. Philips India Ltd. Prodiva 1.5T CS	Responsive	Responsive	-----
3.	M/s.Wipro GE Healthcare Pvt. Ltd GE Signa Creator – 16 channel	Responsive	Responsive	-----

**Sch. 2) Option II-B (10KW/16 channel/ 33mt/m& 120 T/m/s)**

<b>Sl. No</b>	<b>Name of the firm</b>	<b>Technical evaluation report</b>	<b>Commercial evaluation report</b>	<b>Reason for rejection</b>
1	M/s. Philips India Ltd. Prodiva 1.5T CS	Responsive	Responsive	-----
2	M/s.Wipro GE Healthcare Pvt. Ltd GE Signa Creator – 16 channel	Responsive	Responsive	-----

**Sch. 3) Option III A (15KW/16 channel/30 mt/m & 120 T/m/s)**

<b>Sl. No</b>	<b>Name of the firm</b>	<b>Technical evaluation report</b>	<b>Commercial evaluation report</b>	<b>Reason for rejection</b>
1.	M/s. Siemens Healthcare Private Limited MAGNETOM Amira	Responsive	Responsive	-----
2.	M/s. Philips India Ltd. Prodiva 1.5T CS	Non-responsive	Responsive	<p>As per the specialist report, the model quoted by the bidder is not meeting the following point, which is a major deviation. Hence not recommended.</p> <p>1) The model has only 8 channel shoulder coil and knee coil against the tender specification of atleast 16 channel dedicated shoulder and knee and shoulder coil (Sl. No.8 f and g).</p> <p>2. The model has dS SENSE with maximum of 7 elements against the tender specification of Bilateral Breast coil with at least 8 channel with spectroscopy (Sl. No. 8.e)</p>

**Sch. 3) Option III-B (15KW/16 channel/ 33 mt/m & 120 T/m/s)**

<b>Sl. No</b>	<b>Name of the firm</b>	<b>Technical evaluation report</b>	<b>Commercial evaluation report</b>	<b>Reason for rejection</b>
1.	M/s.Wipro GE Healthcare Pvt. Ltd GE Signa Creator – 16 channel	Non-Responsive	Responsive	As per the technical evaluation report, the quoted model is having only 10 kw in digital RF system against the tender requirement of 15kw
2.	M/s.Philips India Ltd PRODIVA	Non-Responsive	Responsive	As per the specialist report, the model quoted by the bidder is not meeting the following point, which is a major deviation. Hence not recommended. 1) The model has only 8 channel shoulder coil and knee coil against the tender specification of atleast 16 channel dedicated shoulder and knee and shoulder coil (Sl. No.8 f and g). 2. The model has dS SENSE with maximum of 7 elements against the tender specification of Bilateral Breast coil with at least 8 channel with spectroscopy (Sl. No. 8.e)

**Sch. 4) Option IV A (18KW/16 channel/ 30mt/m& 120 T/m/s)**

<b>Sl. No</b>	<b>Name of the firm</b>	<b>Technical evaluation report</b>	<b>Commercial evaluation report</b>	<b>Reason for rejection</b>
1	M/s. Siemens Healthcare Private Limited MAGNETOM Altea	Responsive	Responsive	-----
2	M/s.Philips India Ltd MR 5300 with BlueSeal magnet technology	Non- responsive	Responsive	As per the specialist report, the model quoted by the bidder is not

Sl. No	Name of the firm	Technical evaluation report	Commercial evaluation report	Reason for rejection
				meeting the following point, which is a major deviation. Hence not recommended.  1) No integrated and dedicated Knee and shoulder coils available at present.

**Sch. 4) Option IV B (16KW/32 channel/ 44mt/m& 200 T/m/s)**

Sl. No	Name of the firm	Technical evaluation report	Commercial evaluation report	Reason for rejection
1	M/s. Siemens Healthcare Private Limited MAGNETOM SOLA	Responsive	Responsive	-----
2	M/s.Philips India Ltd Ingenia Ambition X 1.5Tesla	Responsive	Responsive	-----
3.	M/s.Wipro GE Healthcare Pvt. Ltd GE Signa Artist (64 channel)	Responsive	Responsive	-----