

## EVALUATION RESULT

**Tender for fixing rate contract for supply and installation of Crash cart,  
Emergency recovery trolley, Semi Fowlers Cot, Spine Board, Scoop  
Board and dressing trolley  
Tender ref. 814/FUR/RC/TNMSC/ENGG/2022, Dt. 12.05.2022.**

-----

### **Schedule no. 1: Crash Cart**

Name of the bidder	Responsiveness		Reasons for rejection if any
	Technical	Commercial	
M/s. Solokrafts Industries, Chennai Model: Kraft 501 Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Imed Furns Pvt. Ltd., Chennai Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Alan Medical Laboratory Products, Chennai Model: AMPL 604 Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Premier Hospital Furniture (Mfg) Co. Chennai Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Prince Hospital Furniture (Mfg) Co., Chennai Model: PHF 039 Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Suraj Meditech Agency, Chennai Model: Suraj crash cart Make: Own manufacturing	Responsive	Non - Responsive	<b>Commercial points:-</b> a) Not furnished EMD as per the terms of the tender. b) Did not furnish details like purchase order, installation report and end user certificate to meet the qualification criteria for past performance as a manufacturer bidder.
M/s. Godrej & Boyce Mfg. Co. Ltd., Chennai Make: Own manufacturing	Non-responsive	Responsive	<b>Technical points:-</b> Lamp and 3 AMP socket not provided
M/s. Lofty Optical Industries, Chennai Make: KM best medical	Non-responsive	Responsive	<b>Technical points:-</b> Five hand out bins and CPR board not provided.
M/s. South India Surgical Co. Ltd., Chennai Make: Own manufacturing	Responsive	Non - Responsive	<b>Commercial points:-</b> Did not furnish details like purchase order, installation report and end user certificate to meet the qualification criteria for past performance as a manufacturer bidder.
M/s. BHN Surgicals Pvt. Ltd., Chennai Model: MW-39 Make: Own manufacturing	Responsive	Non - Responsive	<b>Commercial points:-</b> Did not furnish details like purchase order, installation report and end user certificate to meet the qualification criteria for past performance as a manufacturer bidder.
M/s. Ascend India, Chennai Make: AGI	Responsive	Non - Responsive	<b>Commercial points:-</b> Did not furnish details like purchase order, installation report and end user certificate to meet the qualification criteria for past performance as a manufacturer bidder.
M/s. R. Madhusudhanan, Chennai Make: Own manufacturing	Non -Responsive	Responsive	<b>Technical Points:-</b> a) The joints were not properly welded. b) The welded joints were not properly grinded and was painted with aluminum paint. c) Not sturdy and the looked flimsy.

**Schedule no. 2: Emergency Recovery Trolley**

Name of the bidder with model & make	Responsiveness		Reason for rejection if any
	Technical Evaluation	Commercial Evaluation	
M/s. Solokrafts Industries, Chennai Model: Kraft 400 Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Imed Furns Pvt. Ltd., Chennai Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Premier Hospital Furniture (Mfg) Co. Chennai Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Prince Hospital Furniture (Mfg) Co., Chennai Model: PHF 022A Make: Own manufacturing	Responsive	Responsive	Nil
M/s. South India Surgical Co. Ltd., Chennai Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Ascend India, Chennai Make: AGI	Responsive	Responsive	Nil
M/s. Godrej & Boyce Mfg. Co. Ltd., Chennai Make: Own manufacturing	Non-responsive	Responsive	<b>Technical Points:-</b> Quoted MS trolley against tender specification of should be made of full SS.
M/s. Lofty Optical Industries, Chennai Make: KM best medical	Non-responsive	Responsive	<b>Technical Points:-</b> Quoted MS trolley against tender specification of should be made of full SS.
M/s. BHN Surgicals Pvt. Ltd., Chennai Model: MW-39 Make: Own manufacturing	Non-responsive	Non - Responsive	<b>Technical Points:-</b> Not demonstrated. <b>Commercial points:-</b> Did not furnish details like purchase order, installation report and end user certificate to meet the qualification criteria for past performance as a manufacturer bidder.
M/s. R. Madhusudhanan, Chennai Make: Own manufacturing	Non-responsive	Responsive	<b>Technical Points:-</b> a) The joints were not properly welded. b) The welded joints were not properly grinded and was painted with aluminum paint. c) Not sturdy and the looked flimsy.

**Schedule no. 3: Semi Fowlers Cot**

Name of the bidder with model & make	Responsiveness		Reason for rejection if any
	Technical Evaluation	Commercial Evaluation	
M/s. Godrej & Boyce Mfg. Co. Ltd., Chennai Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Solokrafts Industries, Chennai Model: Kraft 144 Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Lofty Optical Industries, Chennai Make: KM best medical	Responsive	Responsive	Nil
M/s. Imed Furns Pvt. Ltd., Chennai Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Alan Medical Laboratory Products, Chennai Model: AMPL 020 Make: Own manufacturing	Responsive	Responsive	Nil

Name of the bidder with model & make	Responsiveness		Reason for rejection if any
	Technical Evaluation	Commercial Evaluation	
M/s. Premier Hospital Furniture (Mfg) Co. Chennai Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Prince Hospital Furniture (Mfg) Co., Chennai Model: PHF 011 Make: Own manufacturing	Responsive	Responsive	Nil
M/s. South India Surgical Co. Ltd., Chennai Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Ascend India, Chennai Make: AGI	Responsive	Responsive	Nil
M/s. BHN Surgicals Pvt. Ltd., Chennai Model: MW-04-B Make: Own manufacturing	Responsive	Non - Responsive	<b>Commercial points:-</b> Did not furnish details like purchase order, installation report and end user certificate to meet the qualification criteria for past performance as a manufacturer bidder.
M/s. R. Madhusudhanan, Chennai Make: Own manufacturing	Non-responsive	Responsive	<b>Technical Points:-</b> a) The joints were not properly welded. b) The welded joints were not properly grinded and was painted with aluminum paint. c) Not sturdy and the looked flimsy.

**Schedule no. 4: Spine Board**

Name of the bidder with model & make	Responsiveness		Reason for rejection if any
	Technical Evaluation	Commercial Evaluation	
M/s. Thirupathi Enterprises Chennai Model 1: without head mould Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Thirupathi Enterprises Chennai Model 2: with head mould Make: Own manufacturing	Responsive		
M/s. Lofty Optical Industries, Chennai Make: KM best medical	Responsive	Responsive	Nil
M/s. R. Madhusudhanan, Chennai Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Ascend India, Chennai Make: AGI	Responsive	Responsive	Nil
M/s. Prince Hospital Furniture (Mfg) Co., Chennai Model: PHF 01 Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Imed Furns Pvt. Ltd., Chennai Make: Own manufacturing	Non-Responsive	Responsive	Not demonstrated.
M/s. South India Surgical Co. Ltd., Chennai Model: UTS-C1 Make: UT Medical, China	Responsive	Non - Responsive	<b>Commercial points:-</b> Did not furnish details like purchase order, installation report and end user certificate to meet the qualification criteria for past performance as a non-manufacturer bidder.
M/s. BHN Surgicals Pvt. Ltd., Chennai Model: MW-175 Make: Own manufacturing	Responsive	Non - Responsive	<b>Commercial points:-</b> Did not furnish details like purchase order, installation report and end user certificate to meet the qualification criteria for past performance as a manufacturer bidder.

**Schedule no. 5: Scoop Board**

Name of the bidder with model & make	Responsiveness		Reason for rejection if any
	Technical Evaluation	Commercial Evaluation	
M/s. Thirupathi Enterprises Chennai Model 1: Polyethylene Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Prince Hospital Furniture (Mfg) Co., Chennai Model: PHF – P02 Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Thirupathi Enterprises Chennai Model 2: Aluminium Make: Own manufacturing	Non - Responsive	Responsive	<b>Technical points:-</b> Not demonstrated.
M/s. Lofty Optical Industries, Chennai Make: KM best medical	Non - Responsive	Responsive	<b>Technical points:-</b> Quoted aluminium product against the tender specification of polyethylene
M/s. Imed Furns Pvt. Ltd., Chennai Make: Own manufacturing	Non - Responsive	Responsive	<b>Technical points:-</b> Not demonstrated.
M/s. South India Surgical Co. Ltd., Chennai Model: UTS-E4 Make: UT Medical, China	Non - Responsive	Non - Responsive	<b>Technical points:-</b> Quoted aluminum product against the tender specification of polyethylene. <b>Commercial points:-</b> Did not furnish details like purchase order, installation report and end user certificate to meet the qualification criteria for past performance as a non-manufacturer bidder.
M/s. BHN Surgicals Pvt. Ltd., Chennai Model: MW-175 Make: Own manufacturing	Non - Responsive	Non - Responsive	<b>Technical points:-</b> Quoted aluminum product against the tender specification of polyethylene. <b>Commercial points:-</b> Did not furnish details like purchase order, installation report and end user certificate to meet the qualification criteria for past performance as a manufacturer bidder.
M/s. R. Madhusudhanan, Chennai Make: Own manufacturing	Non - Responsive	Responsive	<b>Technical points:-</b> Quoted aluminum product against the tender specification of polyethylene.
M/s. Ascend India, Chennai Make: AGI	Non - Responsive	Responsive	<b>Technical points:-</b> Quoted aluminum product against the tender specification of polyethylene.

**Schedule no. 6: Dressing Trolley**

Name of the bidder	Responsiveness		Reasons for rejection if any
	Technical	Commercial	
M/s. Solokrafts Industries, Chennai Model: Kraft 601 Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Lofty Optical Industries, Chennai Make: KM best medical	Responsive	Responsive	Nil
M/s. Imed Furns Pvt. Ltd., Chennai Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Premier Hospital Furniture (Mfg) Co. Chennai Make: Own manufacturing	Responsive	Responsive	Nil

Name of the bidder	Responsiveness		Reasons for rejection if any
	Technical	Commercial	
M/s. Prince Hospital Furniture (Mfg) Co., Chennai Model: PHF 028 Make: Own manufacturing	Responsive	Responsive	Nil
M/s. South India Surgical Co. Ltd., Chennai Make: Own manufacturing	Responsive	Responsive	Nil
M/s. Ascend India, Chennai Make: AGI	Responsive	Responsive	Nil
M/s. Godrej & Boyce Mfg. Co. Ltd., Chennai Make: Own manufacturing	Non - Responsive	Responsive	<b>Technical points:-</b> Not demonstrated.
M/s. BHN Surgicals Pvt. Ltd., Chennai Model: MW-37 (A) Make: Own manufacturing	Responsive	Non - Responsive	<b>Commercial points:-</b> Did not furnish details like purchase order, installation report and end user certificate to meet the qualification criteria for past performance as a manufacturer bidder.
M/s. R. Madhusudhanan, Chennai Make: Own manufacturing	Non - Responsive	Responsive	<b>Technical Points:-</b> a) The joints were not properly welded. b) The welded joints were not properly grinded and was painted with aluminum paint. c) Not sturdy and the looked flimsy.

**Sd/-**  
**General Manager (E)**