

RESULT OF EVALUATION

TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND INSTALLATION OF ELECTROLYTE ANALYZER, URINE ANALYSER AND ESR AUTOMATIC ANALYZER TO VARIOUS GOVT. INSTITUTIONS

TENDER NO. 571/ANL/RC/TNMSC/ENGG/2020, DT.27.02.2020

Sch.No.1. ELECTROLYTE ANALYZER (Option I, II & III):-

Sl. No	Bidders Quoted	Model/ Make	Responsive bidders		Reason for rejection
			Technical evaluation report	Commercial evaluation report	
1.	M/s. Sensa Core Medical Instrumentation Pvt. Ltd	Option I, II & III: Model: ST200 aQua Own make	Non-responsive	Responsive	As per the specialist report, quoted model does not meet the following technical specification:- 1. Sample volume is 150±30 µL against the tender requirement of equal to or less than 100µL. 2. Primary tube not compatible and require sample cup & pipette tip due to short probe length against the tender requirement of should be compatible to process samples from sample cup, collection tube, capillary etc. 3. 6 hours & 12 hours against the tender requirement of Fully automatic calibration with 2-point calibration every 4 hours and point with every sample
2.	M/s. Transasia Bio Medicals	OPTION I Model: Easylyte	Responsive	Responsive	-----

Sl. No	Bidders Quoted	Model/ Make	Responsive bidders		Reason for rejection
			Technical evaluation report	Commercial evaluation report	
	Ltd	OPTION II Model: Easylyte plus OPTION III Model: Easylyte expand Make – Medical corporation USA			
3.	M/s.Genworks Health Pvt. Ltd.	Option II: Make: M/s. Caretium Medical Instrument Co Ltd Model: Volton Pro	Non-responsive	Non-responsive	Technical points: As per the specialist report, quoted model does not meet the following technical specification:- 1. Required periodic maintenance against the tender requirement of should be provided with long life maintenance free electrodes. 2. Probe cleaning by washing upto 8mm only against the tender requirement of Instrument should have automatic probe wiping facility. Commercial points: Despite given chance, the bidder did not furnish end user certificate for their manufacturer.

Sl. No	Bidders Quoted	Model/ Make	Responsive bidders		Reason for rejection
			Technical evaluation report	Commercial evaluation report	
4.	M/s. Trivitron Healthcare Pvt. Ltd.	Option I & II Model: Elite Make: Biocare corporation			As per the specialist report, quoted model does not meet the following technical specification:- 1. Sample volume is 150 µL against the tender requirement of equal to or less than 100µL. 2. Capillary tube sample not possible against the tender requirement of should be compatible to process samples from sample cup, collection tube, capillary etc.
5.	M/s.AGD Biomedicals Pvt. Ltd	Option III Make: Own Model: EL-120 / AGD3100			As per the specialist report, quoted model does not meet the following technical specification:- 1. Capillary tube sample cannot be processed against the tender requirement of should be compatible to process samples from sample cup, collection tube, capillary etc. 2. Required periodic maintenance against the tender requirement of should be provided with long life maintenance free electrodes.

Sch.No.2. Random Access ESR Automated Analyzer:-

Sl. No	Bidders Quoted	Model/ Make	Responsive bidders		Reason for rejection
			Technical evaluation report	Commercial evaluation report	
1.	M/s. Transasia Bio Medicals Ltd,	Model – Cube 30 touch Make – Diesse diagnostic S.P.A, Italy	Responsive	Responsive	-----

Sch.No.3. URINE ANALYZER:-

Sl. No	Bidders Quoted	Model/ Make	Responsive bidders		Reason for rejection
			Technical evaluation report	Commercial evaluation report	
1.	M/s. Transasia Bio Medicals Ltd,	Model – laura Make – Erba	Responsive	Responsive	-----
2.	M/s.M.R. Enterprises	U-500 Acon, mission	Responsive	Non-responsive	Despite given chance, the bidder did not furnish end user certificate for their manufacturer and also for them as non-manufacturer bidder and also not furnished cash-credit limit certificate from banker.
3.	M/s. Sun Medica,	Make: Siemens Clinitek- status+ analyser	Substantially Responsive	Responsive	Remarks:- The deviations are stated as minor by the specialist.