

EVALUATION RESULT

Tender for fixing rate contract for supply and installation of Flowcytometer

Tender ref. no. 887/FCY/RC/TANSAC/TNMSC/ENGG/2022, dt. 30.11.2022

Name of the bidder with model & make	Responsiveness		Reason for rejection if any
	Technical Evaluation	Commercial Evaluation	
M/s. Navi Medicare Pvt. Ltd., Chennai Model: BD FACSLytic Make: BD and Company	Responsive	Responsive	Nil
M/s. Beckman Coulter India Pvt. Ltd., Model: DxFlex System B4-R2-V0 Make: Own	Non-Responsive	Responsive	1) Not US FDA certified. Only CFDA certified. 2) Only area and height parameter is available against tender specification of area, height and width parameter. 3) PMT not available.

Sd/-
General Manager (E)