

27.01.2023

**Tender For Supply and Installation of Ultrasonograph for Govt. Hospitals
at Avadi – Tiruvallur District, Velampalayam – Tirpuur District, Ammapet
– Salem district and Kandiaperi – Tirunelveli District**

Tender No. PKG3/C3/ICB/TNUHP/JICA/TNMSC/ENGG/2022,

Dt. 13.07.2022

Result of Evaluation

Sl. No	Bidders Quoted	Equipment Details	Responsiveness		Reason for rejection / Remarks
			Technical Responsiveness	Commercial Responsiveness	
1. Mobile Ultrasound Machine – 5 Nos.					
1.	M/s. Esaote India	Make: Esaote S.p.A, Italy Model: MyLab Omega	Responsive	Responsive	-----
2.	M/s. Solokrafts Industries	Make: Mindray Model: M6	Non - Responsive	Responsive	Based on Specialist remarks, the quoted model does not meet the following tender specification: System should have two or more active probe connectors and probe holders in the main unit. – <i>Only one connector is available in the main unit</i>
3.	M/s. Wipro GE	Make: GE Model: Versana Active	Non - Responsive	Responsive	Based on Specialist remarks, the quoted model does not meet the following tender specification: System should have two or more active probe connectors and probe holders in the main unit. – <i>Only one connector is available in the main unit</i>
2. Ultrasound Machine – 4 Nos.					
1	M/s. Solokrafts Industries	Make: Mindray Model: Resona I9	Responsive	Responsive	-----

Sl. No	Bidders Quoted	Equipment Details	Responsiveness		Reason for rejection / Remarks
			Technical Responsiveness	Commercial Responsiveness	
2	M/s. Esaote India	Make: Esaote S.p.A, Italy Model: MyLab X8	Responsive	Responsive	-----
3	M/s. Wipro GE	Make: GE Model: Logiq P9	Responsive	Responsive	-----

Sd/-

General Manager(E)