

Date: 03.02.2023

**TENDER FOR SUPPLY AND INSTALLATION OF ENT EQUIPMENT-I FOR
TERTIARY CARE HOSPITALS AT GOVT. RAJAJI HOSPITAL, MADURAI,
AND GOVT. KILPAUK MEDICAL COLLEGE HOSPITAL, CHENNAI IN
TAMIL NADU**

PKG14/C1/ICB/TNUHP/ JICA/TNMSC/ENGG/ 2022, dt. 29.07.2022

RESULT OF EVALUATION

Sl. No.	Name of the bidder and Equipment Details	Responsiveness		Final Responsiveness / Remarks
		Technical	Commercial	
CO2 Laser for ENT (Lot/Contract no.1)				
1.	M/s Boston Scientific India Pvt. Ltd., Gurgaon Make: Lumenis, Israel Model: Acupulse Duo a) <u>Micromanipulator:</u> Digital Acublade b) <u>Fiber delivery:</u> Flexible laser fibers	Responsive	Substantially Responsive	Substantially Responsive
2.	M/s Vaansari Marketing Services, Chennai Make: Vaansari Marketing Services (VMS), India Model: Energi a) <u>Micromanipulator:</u> TTI Medical, USA b) <u>Fiber delivery:</u> Reusable fiber	Non-Responsive	---	Technical points: A. Micromanipulator with following requirements for micro laryngeal laser surgery: k) It should have a robotic laser microsurgery system with following requirements: (ii)It should have a penetration depth of 0.2 mm to 2 mm. (user defined): Not mentioned B.Oral, Pharyngeal and nasal hand piece set which includes:

Sl. No.	Name of the bidder and Equipment Details	Responsiveness		Final Responsiveness / Remarks
		Technical	Commercial	
				<p>i)230mm hand piece unit (cvd optical unit, ports holder, m conical main extender, contamination collector), extra conical main extender, backstop extender - 3 no's, tip extender - 3nos, straight tip, kamami nasal tip - 3 Nos, kamami tonsil tip - 3 Nos, 90-degree angled mirror tip extender, cleaning brush, tygon tube (8mm id, 1.5m long) w/reducer fitting: Not mentioned.</p> <p>C. Fiber Accessories:</p> <p>a. Reusable CO2 fiber - 01 no: Numbers not mentioned</p> <p>b. Rigid hand piece kit at least 8 rigid hand pieces: Numbers not mentioned</p> <p>(i)60mm, straight, straight tip, 180mm, straight, straight tip, 60mm, straight, curved tip, 140mm, straight, curved tip 180mm, straight, curved tip, 240mm, bent, curved tip, 140mm, bent, straight</p>

Sl. No.	Name of the bidder and Equipment Details	Responsiveness		Final Responsiveness / Remarks
		Technical	Commercial	
				<p>tip, 240mm, bent, straight tip: Not available in product catalogue.</p> <p>c. Endoscope protection sheath - 2 no's length: 640 mm, OD: 1.7mm: Not available in product catalogue.</p> <p>d. Hand piece bending tool: Not available in product catalogue.</p> <p>e. Hand piece cleaning kit: includes 3 cleaning brushes and 20 extra silicone tubes for hand pieces: Not available in product catalogue.</p> <p>f. Bending and cutting tools to reuse fiber: Not available in product catalogue.</p> <p>g. Sterilization tray for fibers: Not available in product catalogue.</p> <p>D. Smoke Evacuator: Compatible with the laser machine, imported quality – includes: Smoke evacuation unit with pneumatic footswitch, via 6 filter-6</p>

Sl. No.	Name of the bidder and Equipment Details	Responsiveness		Final Responsiveness / Remarks
		Technical	Commercial	
				<p>hour double port 7/8" and 1 1/4", 7/8" tubing with wand and tip-2 no's, sml of 50-laser mask 0.1 mm filtration media(flat mask)laser mask 0.1 mm filtration media (flat mask): Not available in product catalogue.</p> <p>E. Other requirements:</p> <p>b. There should be a replacement machine in case the machine goes out of function within 72 hours: Not available in product catalogue.</p> <p>From the above, in the technical evaluation report furnished by the Expert on 24.11.2022, had requested the firm to furnish the data sheet for their quoted product. Accordingly, in email dt. 28.11.2022 the firm was requested to furnish their clarifications on the above points on or before 30.11.2022</p>

Sl. No.	Name of the bidder and Equipment Details	Responsiveness		Final Responsiveness / Remarks
		Technical	Commercial	
				<p>However, the firm in turn didn't furnish their reply for this subject.</p> <p>Hence, their bid is considered as "Non-Responsive".</p> <p><u>Commercial points:</u> The bid of M/s Vaansari Marketing Services, Chennai was not considered for commercial evaluation, as their model was rejected by the Specialists in Technical Evaluation.</p>
Nasal Endoscopy complete set (Lot/Contract no.2)				
1.	<p>M/s Vishal Surgical Equipment Company Private Limited, Chennai</p> <p>Make: Karlstorz, Germany Model:</p> <p>a) <u>Full 4K HD video image processor:</u> Image 1S Connect II, Connect Module, Image 1 S 4U-Link, link module</p> <p>b) <u>Full 4K Camera Head:</u> Image 1S 4U Rubina, OPAL 1® NIR/ICG 2 chip 4K UHD camera head</p> <p>c) <u>LED light source:</u> Cold light fountain power LED Rubina</p> <p>d) <u>Light guide cable:</u> 495TIP fiberoptic light cable</p> <p>e) <u>55" (or more) Monitor:</u> 58" 4K monitor Vesa 400x400 and Vesa 400x200</p>	Responsive	Responsive	Responsive

Sl. No.	Name of the bidder and Equipment Details	Responsiveness		Final Responsiveness / Remarks
		Technical	Commercial	
	f) Telescopes: Hopkins 0 deg, 30 deg, 45 deg and 70 deg (4mm dia and length 18cm- each) g) <u>Telescopes handles-</u> Stammberger [flat- 723770 and round- 723771] h) <u>Trolley:</u> Indian Make			

Sd/-

General Manager (E)