

Date: 16.02.2023

**TENDER FOR SUPPLY AND INSTALLATION OF MAGNETIC
RESONANCE IMAGING SYSTEM (MRI)
FOR TERTIARY CARE HOSPITALS AT GOVT. KILPAUK
MEDICAL COLLEGE HOSPITAL, CHENNAI AND
GOVT. COIMBATORE MEDICAL COLLEGE HOSPITAL,
COIMBATORE.**

**PKG20/C1/ICB/TNUHP/ JICA/TNMSC/ENGG/ 2022, dated 29.07.2022
RESULT OF EVALUATION**

Sl. No	Bidders Quoted	Equipment Details	Responsiveness		Remarks
			Technical	Commercial	
Magnetic Resonance Imaging System (MRI)					
1.	M/s Philips Medical Systems, Netherlands B.V.,	Make: Philips, Nederland Model: Ingenia Elition X 3T MRI Scanner	Substantially Responsive	Responsive	<u>Reasons for Substantial Responsiveness:</u> a) Against the requirement of RF Coils 16 channel Head Coil or more head coil: dStream Head Coil with 15 Channel offered. b) Against the requirement of Loop flex coils 16 channels or more (large and small): Loop Flex Coils Small (4ch), Medium (6ch), Large(8ch) with 16 Channel Small Extremity coils are offered.

Sl. No	Bidders Quoted	Equipment Details	Responsiveness		Remarks
			Technical	Commercial	
2.	M/s Siemens Healthcare Pvt. Ltd., India	Make: Siemens Healthcare, Germany Model: Magnetom Vida syngo MR XA50	Substantially Responsive	Responsive	<p><u>Reasons for Substantial Responsiveness:</u></p> <p>a) The requirement of 10 nos. of disposable endo coils for for rectal and vaginal insertions:Not offered.</p> <p>b) Against the requirement of 16 channels or more Dedicated RIGID coil-Shoulder, Wrist, Foot, Knee, Elbow, Ankle: Offered following coils as standard - Shoulder Shape 16 #3T ; Tx/Rx Knee 18 #3T; Hand/Wrist 16 #3T and Foot/Ankle 16 #3T are offered. For elbow coil</p> <p><u>Tx/Rx coil- 18 channel coil (rigid coil) and</u> Ultraflex 18 large and small are offered for elbow imaging.</p>

Sl. No	Bidders Quoted	Equipment Details	Responsiveness		Remarks
			Technical	Commercial	
					c) Against the requirement of Nordic or any FDA approved Make for fMRI package - FDA approved Kryptonite is accepted.
3.	M/s Wipro GE Healthcare Private Limited, Singapore	Make: GE Healthcare, Singapore Model: Envision Signa Pioneer	Non-Responsive	Due to non-responsiveness on technical specifications, commercial points are not evaluated.	<u>Reasons for rejection:</u> a) Against the requirement of Activity shielded Gradient System with strength of at least 45 mT / m, offered strength is only 36 mT/m. b) Against the requirement of Slew rate of the system should be at least 200 T/ m /sec: the offered strength is only 150T/ m /sec mentioned.

Sd/-
General Manager (E)