

23.03.2023

International Competitive Bidding (ICB) For Supply And Installation Of PCA Pump and Blood/ Fluid Warmer For Tertiary Care Hospitals At Govt. Rajaji Hospital – Madurai, Govt. Kilpauk Medical College Hospital – Chennai And Govt. Coimbatore Medical College Hospital - Coimbatore

IFB No.PKG5/C1/ICB/TNUHP/JICA/TNMSC/ENGG/2022 dt 29.07.2022

Result of Evaluation

Sl. No	Bidders Quoted with make and model	Responsiveness		Reason for rejection / Remarks
		Technical Responsiveness	Commercial Responsiveness	
Patient Controlled Analgesia (PCA) Pump – 22 Nos.				
1.	M/s. Miti Electronics Make: Fresinius Kabi Model: Agilia SP PCA	Responsive	Responsive	-----
Blood and Fluid Warmer – 79 Nos.				
1	M/s. Madras Surgicals and Chemicals Make: Smiths Model: LEVEL 1 Hotline	Responsive	Responsive	-----
2	M/s. Indian Surgical Equipment Co. Make: Tahat Model: Ampir - 01	Non - Responsive	Non - Responsive	Technical Points: Based on the specialist remarks, the quoted model does not meet the following tender specifications: It should be operating in triple lumen tubing to maintain layer of 42 ° C circulating solution around IV line. Commercial points: Did not furnish the following documents: Purchase order and performance satisfactory certificate meeting the minimum tender requirement of 47 Nos. for

Sl. No	Bidders Quoted with make and model	Responsiveness		Reason for rejection / Remarks
		Technical Responsiveness	Commercial Responsiveness	
				<p>Manufacturer and 24 nos. for Non – Manufacturer</p> <p>Did not furnish the following forms for Manufacturer:</p> <ol style="list-style-type: none"> 1. FIN- : Average Annual Turnover 2. FIN -3: Financial Resources 3. EXP -1: General Experience 4. EXP-2: Specific Experience

Sd/-
General Manager(E)