

05.05.2023

**Online Tender For the supply and Installation of Anesthesia Workstation
(MRI Compatible, 2 and 3 gas System) to Three Tertiary Care Hospital at
Govt. Rajaji Hospital, Madurai, Govt. Kilpauk Medical College Hospital,
Chennai and Govt. Medical College Hospital, Coimbatore**

**Ref: Tender No.PKG7/C1/ICB/TNUHP/JICA/TNMSC/ENGG/2022, Dt.
29.07.2022**

RESULT OF EVALUATION

Sl. No.	Name of Bidder with make and model	Responsiveness		Remarks/Reason for rejection
		Technical Responsiveness	Commercial Responsiveness	
I	Anesthesia Workstation (3 Gas System)– 47 Nos.			
1	M/s. Madras Surgicals and Chemicals Make: Draeger Model: Fabius Plus	Responsive	Responsive	-----
2	M/s. Solokrafts Industries Make: Mindray Model: Wato – Ex 35	Non - Responsive	Responsive	As per specialists remarks, the quoted model does not meet the following point: In case of Oxygen supply failure, shall be able to continue ventilation (without interruption) by automatically switching over to atmospheric air through integrated ventilator from same manufacturer or any other technology for safety of the patients, which should be authorized and advised and properly documented by manufacturer with all quality standard certificates & practically demonstrated as per ISO 80601-2-13 regulations without delivery of Anaesthesia Agent. – <i>Ventilation stops during oxygen supply failure and resumes only after manual interruption. Patient safety is compromised.</i>
II	Anesthesia Workstation (2 Gas System) – 7 Nos.			
1	M/s. Solokrafts Industries	Responsive	Responsive	-----

Sl. No.	Name of Bidder with make and model	Responsiveness		Remarks/Reason for rejection
		Technical Responsiveness	Commercial Responsiveness	
	Make: Mindray Model: Wato – Ex 30			
III	Anesthesia Workstation (MRI compatible) – 2 Nos.			
1	M/s. Madras Surgicals and Chemicals Make: Draeger Model: Fabius MRI	Responsive	Responsive	-----

Sd/-
General Manager(E)