

08.05.2023

**International Competitive Bidding (ICB) For the Supply and Installation of
Ultrasound System - II to Three Tertiary Care Hospitals at Govt. Rajaji
Hospital, Madurai, Govt. Kilpauk Medical College Hospital, Chennai and Govt.
Coimbatore Medical College Hospital, Coimbatore**

Tender No.: PKG9/C1/ICB/TNUHP/JICA/TNMSC/ENGG/2022, Dt. 29.07.2022

Result of Evaluation

Sl. No.	Name of Bidder with quoted make and model	Responsiveness		Remarks/ Reasons for Rejection
		Technical Responsiveness	Commercial Responsiveness	
I.	Ultrasound machine for Radiology – 2 Nos.			
1	M/s. Esaote India (NS) Ltd. Make: Esaote S.p.A Italy Model: MyLab X8 eXP	Responsive	Responsive	-----
2	M/s. Solokrafts Industries Make: Mindray Model: Resona I9	Substantially Responsive	Responsive	-----
II	Ultrasound Machine for Radiology with 4D Probe and Elastography – 3 Nos.			
1	M/s. Esaote India (NS) Ltd. Make: Esaote S.p.A Italy Model: MyLab X8 eXP	Responsive	Responsive	-----
2	M/s. Solokrafts Industries Make: Mindray Model: Resona I9	Non - Responsive	Responsive	As per the Specialist evaluation report, the quoted model does not meet the following specifications: i. System should have Automatic IT measurement – <i>Not Available</i> ii. System should have Hyperechogenic Spot structure detection and Enhancement to detect and Highlight Calcifications in Breast - <i>Not Available</i> iii. System should have SWE both in Convex as well as High and low frequency linear probes – <i>SWE not available</i>

Sl. No.	Name of Bidder with quoted make and model	Responsiveness		Remarks/ Reasons for Rejection
		Technical Responsiveness	Commercial Responsiveness	
				iv. Shear Wave not available in the following transducers: a. Convex array probe with frequency range of 1-8 MHz, the probe should have a pin-less connector. b. Linear array probe with frequency range of 4-15 MHz, the probe should have a pin-less connector. c. Linear array probe with frequency range of 3- 11 MHz, the probe should have a pin-less connector.
III	Ultrasound machine (Nerve Block) – 9 Nos.			
1	M/s. Esaote India (NS) Ltd. Make: Esaote S.p.A Italy Model: MyLab Sigma Elite	Responsive	Responsive	-----
2	M/s. Solokrafts Industries Make: Mindray Model: MX7	Non - Responsive	Responsive	As per the Specialist evaluation report, the quoted model does not meet the following specifications: i. Should have Tissue Enhancement Imaging – <i>Not available</i>
IV	Ultrasound Machine(Portable) – 4 Nos.			
1	M/s. Esaote India (NS) Ltd. Make: Esaote S.p.A Italy Model: MyLab Sigma	Responsive	Responsive	-----
2	M/s. Solokrafts Industries Make: Mindray Model: MX7	Substantially Responsive	Responsive	-----
V	Ultrasound machine with doppler and all probes – 1 No.			
1	M/s. Focus Imaging Systems Make: Philips Model: Affinitti 30	Substantially Responsive	Responsive	-----

Sl. No.	Name of Bidder with quoted make and model	Responsiveness		Remarks/ Reasons for Rejection
		Technical Responsiveness	Commercial Responsiveness	
2	M/s. Solokrafts Industries Make: Mindray Model: DC 40 Full HD	Non - Responsive	Responsive	As per the Specialist evaluation report, the quoted model does not meet the following specifications: i. Cine loop as well as cine scroll facility in B mode with storage of 1000 or more images should be available – <i>Not available</i> ii. Should have Real Time Compound Imaging Technology with Multiple transmitted lines of sight in convex, linear and Endo cavity probes. – <i>Not available in endocavity probe.</i> iii. Should be offered with Speckle Reduction Imaging, panoramic imaging should be available. – <i>Panaromic imaging not available</i>
VI	Ultrasound with doppler for Dialysis ICU – 1 No.			
1	M/s. Esaote India (NS) Ltd. Make: Esaote S.p.A Italy Model: MyLab Sigma	Responsive	Responsive	-----
2	M/s. Solokrafts Industries Make: Mindray Model: MX7	Non - Responsive	Responsive	As per the Specialist evaluation report, the quoted model does not meet the following specifications: i. Tissue Enhancement Imaging – <i>Not available</i> ii. Linear array “Hockey Stick Style” probes of frequency 6 -18 MHZ tolerance

Sl. No.	Name of Bidder with quoted make and model	Responsiveness		Remarks/ Reasons for Rejection
		Technical Responsiveness	Commercial Responsiveness	
				of +/- MHz with connector – <i>Not available in data sheet.</i>
VII	Ultrasound machine with Transcranial Doppler – 1 No.			
1	M/s. Esaote India (NS) Ltd. Make: Esaote S.p.A Italy Model: MyLab X5	Non - Responsive	Responsive	As per the Specialist evaluation report, the quoted model does not meet the following specifications: i. System must be offered with frequency compounding facility. Other equivalent technology can also be offered. – <i>Not available</i> ii. Should have state of the art Transmit Real Time Compound Imaging Technology – <i>Not available</i>
2	M/s. Solokrafts Industries Make: Mindray Model: DC 40 Full HD	Responsive	Responsive	-----

Sd/-
General Manager(E)