

12.06.2023

**International Competitive Bidding (ICB) For the Supply and Installation
of Bronchoscope and Nasopharyngoscope to Three Tertiary Care
Hospitals at Govt. Rajaji Hospital, Madurai, Govt. Kilpauk Medical
College Hospital, Chennai and Govt. Coimbatore Medical College
Hospital, Coimbatore**

**Tender No: PKG11/C1/ICB/TNUHP/JICA/TNMSC/ENGG/2022, Dt.
29.07.2022**

RESULT OF EVALUATION

Sl. No.	Name of the bidder with make and model	Responsiveness		Remarks/ Reason for rejection
		Technical Responsiveness	Commercial Responsiveness	
I	Nasopharyngoscope – No bids received			
II	Intubating Bronchoscope (Fibre Optic)			
1	<p>M/s. Olympus Medical systems Make: Olympus, Japan Model: <u>For Adult:</u> 1. Tracheal Intubation Fiberscope: LF-TP, 2. LED Light source: CLL-V1, 3. Light Guide Cable: WA03310A, 4. Light source: WA91502A, 5. Leakage Tester: WA23080A <u>For Pediatric:</u> Tracheal Intubation Fiberscope: LF-DP, 2. LED Light source: CLL-V1, 3. Light Guide Cable: WA03310A, 4. Light source: WA91502A, 5. Leakage Tester: WA23080A</p>	Non- Responsive	Non- Responsive	<p>As per the Specialist evaluation report, the quoted model does not meet the following tender specification: Specification for Fibreoptic Bronchoscope – Paediatric i. Against the requirement of Insertion tube should have maximum 2.8mm diameter the quoted model is with diameter 3.5 mm ii. Against the requirement of Distal end should have maximum 2.6 mm diameter the quoted model is with diameter 3.1 mm. Commercial terms: The no. of purchase order and their relevant performance certificates to be furnished does not meet the required qualification criteria.</p>
2	<p>M/s. South India Surgical Co. Ltd Make: Pentax Model: For Adult: FI-16RBS</p>	Responsive	Non- Responsive	<p>Commercial terms: The bidder has not furnished purchase order and their relevant performance certificates for the supply of same</p>

Sl. No.	Name of the bidder with make and model	Responsiveness		Remarks/ Reason for rejection
		Technical Responsiveness	Commercial Responsiveness	
	For Pediatric: FB - 8V			make and similar model in the last 5 years as per the requirement if the tender.
III	Flexible Video Bronchoscope			
1	M/s. Fujifilm Make: Fujifilm, Japan Model: EB-530T, VP-3500HD, XL-4450	Responsive	Responsive	-----
2	M/s. Olympus Medical systems Make: Olympus, Japan Model: 1. Video system: CV-170 2. Bronchovideoscope: BF-1TQ170 3. Monitor: LMD – 2435MD 4. Leakage Tester: MB-155	Responsive	Responsive	-----
3	M/s. South India Surgical Co. Ltd Make: Pentax Model: Video Processor: EPK-3000 HD Bronchovideoscope: EB-19J10, Leakage Tester: SHA-p5, Monitor: LMD-2451MD	Responsive	Non- Responsive	Commercial terms: The bidder has not furnished purchase order and their relevant performance certificates for the supply of same make and similar model in the last 5 years as per the requirement if the tender.
IV	Rigid Bronchoscope with Optical Forceps pediatric			
1	M/s. Vishal Surgical Equipment Make: Karl Storz Model: Telescope: Hopkins	Responsive	Responsive	-----

Sd/-

General Manager(E)