

26.06.2023

**International Competitive Bidding (ICB) For the Supply and Installation of  
Ultrasound System - II to Three Tertiary Care Hospitals at Govt. Rajaji  
Hospital, Madurai, Govt. Kilpauk Medical College Hospital, Chennai and Govt.  
Coimbatore Medical College Hospital, Coimbatore**

**Tender No.: PKG9/C1/ICB/TNUHP/JICA/TNMSC/ENGG/2022, Dt. 29.07.2022**

**Result of Re - Evaluation**

Sl. No.	Name of Bidder with quoted make and model	Responsiveness		Remarks/ Reasons for Rejection
		Technical Responsiveness	Commercial Responsiveness	
<b>I.</b>	<b>Ultrasound machine for Radiology – 2 Nos.</b>			
1	M/s. Esaote India (NS) Ltd. Make: Esaote S.p.A Italy Model: MyLab X8 eXP	Responsive	Responsive	-----
2	M/s. Solokrafts Industries Make: Mindray Model: Resona I9	Substantially Responsive	Responsive	-----
<b>II</b>	<b>Ultrasound Machine for Radiology with 4D Probe and Elastography – 3 Nos.</b>			
1	M/s. Esaote India (NS) Ltd. Make: Esaote S.p.A Italy Model: MyLab X8 eXP	Responsive	Responsive	-----
2	M/s. Solokrafts Industries Make: Mindray Model: Resona I9	Non - Responsive	Responsive	As per the specialist report, the quoted model does not meet the following specifications: i. System should have Automatic IT measurement - <i>Smart NT stated as available with a claim that it can perform both NT and IT cannot be accepted.</i> ii. System should have Hyperechogenic Spot structure detection and Enhancement to detect and Highlight Calcifications in Breast - <i>The S/W SmartBreast indicated in the datasheet does not explicitly evidence the possibility of detecting Hyperechogenic Spot structure detection and Enhancement to detect and Highlight Calcifications in Breast</i>

Sl. No.	Name of Bidder with quoted make and model	Responsiveness		Remarks/ Reasons for Rejection
		Technical Responsiveness	Commercial Responsiveness	
				<i>required as per tender specification and no additional documentary evidence is provided for the above. Also, the video evidence clips provided shows only mass characteristics and not calcifications.</i>
<b>II I</b>	<b>Ultrasound machine (Nerve Block) – 9 Nos.</b>			
1	M/s. Esaote India (NS) Ltd. Make: Esaote S.p.A Italy Model: MyLab Sigma Elite	Responsive	Responsive	-----
2	M/s. Solokrafts Industries Make: Mindray Model: MX7	Non - Responsive	Responsive	As per the specialist report, the quoted model does not meet the following: i. Should have Tissue Enhancement Imaging - <i>The HDScope claimed as available is not meeting the requirement of Anesthetists for performing Nerve Block.</i>
<b>IV</b>	<b>Ultrasound Machine(Portable) – 4 Nos.</b>			
1	M/s. Esaote India (NS) Ltd. Make: Esaote S.p.A Italy Model: MyLab Sigma	Responsive	Responsive	-----
2	M/s. Solokrafts Industries Make: Mindray Model: MX7	Substantially Responsive	Responsive	-----
<b>V</b>	<b>Ultrasound machine with doppler and all probes – 1 No.</b>			
1	M/s. Focus Imaging Systems Make: Philips Model: Affiniti 30	Substantially Responsive	Responsive	-----
2	M/s. Solokrafts Industries Make: Mindray Model: DC 40 Full HD	Non - Responsive	Responsive	As per the specialist report, the quoted model does not meet the following: i. Cine loop as well as cine scroll facility in B mode with storage of 1000 or more images should be available - <i>The storage of 1000</i>

Sl. No.	Name of Bidder with quoted make and model	Responsiveness		Remarks/ Reasons for Rejection
		Technical Responsiveness	Commercial Responsiveness	
				<i>images cannot be claimed as equivalent to the 480 seconds cine loop and frame rate indicated in the data sheet.</i>
<b>VI</b>	<b>Ultrasound with doppler for Dialysis ICU – 1 No.</b>			
1	M/s. Esaote India (NS) Ltd. Make: Esaote S.p.A Italy Model: MyLab Sigma	Responsive	Responsive	-----
2	M/s. Solokrafts Industries Make: Mindray Model: MX7	Non - Responsive	Responsive	As per the specialist report, the quoted model does not meet the following: Should have Tissue Enhancement Imaging - <i>The HD Scope claimed as available is not meeting the requirement of Nephrologists for performing dialysis procedure.</i>
<b>VI I</b>	<b>Ultrasound machine with Transcranial Doppler – 1 No.</b>			
1	M/s. Esaote India (NS) Ltd. Make: Esaote S.p.A Italy Model: MyLab X5	Non - Responsive	Responsive	As per the Specialist evaluation report, the quoted model does not meet the following specifications: i. System must be offered with frequency compounding facility. Other equivalent technology can also be offered. – <i>Not available</i> ii. Should have state of the art Transmit Real Time Compound Imaging Technology – <i>Not available</i>
2	M/s. Solokrafts Industries Make: Mindray Model: DC 40 Full HD	Responsive	Responsive	-----

Sd/-

**General Manager(E)**