

30.08.2023

**ONLINE TENDER FOR SUPPLY AND INSTALLATION OF OPHTHALMOLOGY  
EQUIPMENT-I TO THREE TERTIARY CARE HOSPITAL AT GOVT. RAJAJI  
HOSPITAL, MADURAI; GOVT KILPAUK MEDICAL COLLEGE HOSPITAL,  
CHENNAI AND GOVT. COIMBATORE MEDICAL COLLEGE HOSPITAL,  
COIMBATORE**

**PKG62/C1.2/P2/LCB/TNUHP/JICA/TNMSC/ENGG/2023, Dt. 09.03.2023**

**RESULT OF EVALUATION**

<b>Sch. No.</b>	<b>Name of the Bidder</b>	<b>Responsiveness on Technical Points</b>	<b>Responsiveness on Commercial Terms</b>	<b>Reason for rejection/ Remarks</b>
<b>I</b>	<b>Specular Microscope (in vivo)</b>			
1.	No Bids Received	-----	-----	-----
<b>II</b>	<b>Surgical Operating Microscope for advanced Vitreo Retinal Surgery</b>			
1.	<b>M/s. Appasamy Associates Pvt. Ltd.</b> Make: APPASAMY ASSOCIATES PVT. LTD. Model: LEGEND TRUGLOW FS-1	Non Responsive	Responsive	As per specialist evaluation report, the quoted model does not have the following specification:- 8) Xenon short arc reflector lamp with Ha Mode filter. 11) 25 % gray filter Blue blocking filter Fluorescence filter 12) Slit widths: 0.2 mm, 2 mm, 3 mm, 4 mm Hence <b>not recommended</b>
2.	<b>M/s. Care and Care Medical Systems</b> Make: CARL ZEISS, GERMANY Model: OPMI LUMERA 700	Responsive	Responsive	-----
<b>III</b>	<b>Confocal Microscopy</b>			
1.	No Bids Received	-----	-----	-----

**Sd/-**

**General Manager(E)(i/c)**