

13.09.2023

**Online Tender for Supply and Installation of Intravascular Ultrasound
System with Fractional Flow Reserve System to Kalaignar Centenary
Super Speciality Hospital in The Premises of King Institute of
Preventive Medicine and Research Guindy, Chennai
Online tender ref. E969/IVUS/KMSSH/TNMSC/ENGG/2023,
Dt.24.04.2023**

TECHNICAL EVALUATION REPORT

S. No	Name of the bidder	Responsiveness of bidders		Reason for Rejection
		Technical Evaluation report	Commercial evaluation report	
1	M/s. Veiva Scientific India Pvt. Ltd., Model: Intravision 5 Mobile IVUS/ FFR/IFR System Make: Philips	Non- responsive	Not evaluated, Due to Non- responsiveness in technical evaluation.	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. Hence not recommended. IVUS: <input type="checkbox"/> The modality supplies of Intracardiac (ICE) is not available. <input type="checkbox"/> The compatibility of coronary (60MHz High Definition & 40MHz) catheter and Intra cardiac Echo (ICE-(30MHz) Catheter are not available. <input type="checkbox"/> 60MHz High-Definition catheter is not available. FFR System: <input type="checkbox"/> Wireless Bluetooth Connectivity is not available. <input type="checkbox"/> Offered model has IFR (Instantaneous Wave Free Ratio) instead of DFR (Diastolic Hyperemia-Free Ratio).
2	M/s. Boston Scientific India Pvt.Ltd., Model: iLab Polaris Multi-Modality Guidance system with Fractional flow reserve systems. Make: Boston Scientific	Responsive	Responsive	-----

**Sd/-
General Manager (E)i/c**