

Dt.12.09.2023

EVALUATION REPORT

**Online Tender for Fixing Rate contract for supply and installation of
Centrifuge, Refrigerated Centrifuge and Centrifuge-Serofuge**

Tender no. E990/CEN/RC/TNMSC/ENGG/2023, Dt.08.05.2023

Sch.no.1 Centrifuge

Sl. No	Name of the Bidder	Responsiveness of bidder		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Agile Lifescience Technologies India Pvt. Ltd, Make: Andreas Hettich, Germany, Model: Rotofix 32 A	Responsive	Responsive	-----
2	M/s. Alan Medical & Laboratory Products, Make: Remi, Model: R8M Plus with S 4-100, B 100, B 9-5, RA 166, AR 6-50, RA 164	Non-responsive	Responsive	The model quoted did not meeting the following technical points. <ul style="list-style-type: none">• The centrifuge should be capable of accepting a 12 slide cyto rotor, and must be clearly mentioned with details of the cyto rotor system – not available.• Capacity – Accommodate from 133 tubes of 0.5ml to 4 tubes of 100 ml. Adapters should be available for tubes of volume range between 0.5 ml to 100 ml including vaccutainers. Clinical labs have many tubes for processing and versatile centrifuge is preferable not available.

Sch.no.2 Refrigerated Centrifuge

Sl. No	Name of the Bidder	Responsiveness of bidder		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Agile Lifescience Technologies India Pvt. Ltd, Make: Andreas Hettich, Germany, Model: Rotosilenta 630 RS	Responsive	Responsive	-----
2	M/s. Alan Medical & Laboratory Products, Make: Remi, Model: KBM80 Plus with K751, K753, PB753M, PB753L, PB753BF, VS07	Non-responsive	Responsive	The quoted model not meeting the following technical points:- <ul style="list-style-type: none"> • Speed Control Range : 500 - 4500RPM – <u>Offered maximum range 4200RPM & minimum control range not mentioned.</u> • Timer: 0-99 minutes with minimum resolution of 1 minute. 1 sec-999 min: 59 sec and continuous run is also acceptable – <u>Offered 0-99 hrs 59 mins, seconds not mentioned.</u> • Design: Stable, Study all steel design with stainless steel rotor chamber. Easy to clean/corrosion resistant painting – <u>Offered PUF insulated crossion free stainless armoured chamber; Rotor has to be removed for cleaning.</u>

Sch.no.3 Centrifuge-Serofuge

Sl. No	Name of the Bidder	Responsiveness of bidder		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Alan Medical & Laboratory Products, Make: Remi, Model: R8C Plus with A 24-15	Responsive	Responsive	-----