

26.09.2023

**Online Tender for Supply and Installation of Refrigerator for Skin Graft Storage
To Three Tertiary Care Hospital At Govt. Kilpauk Medical College Hospital,
Chennai and Govt. Coimbatore Medical College Hospital, Coimbatore**

Quotation No. PKG35(R1)/C1/LT/TNUHP/JICA/TNMSC/ENGG/2023,

Dt. 26.07.2023

Result of Evaluation

Sl. No	Name of the bidder with make and model	Position	Technical responsiveness	Reasons for rejection
1.	M/s. Vin Bio Care Make: Suzhou Being Medical Device Model: BDW-86L390	L1	Responsive	-----
2.	M/s. Alan Medical And Laboratory Products Make: Remi Model: ULT360 With Stabilizer	L2	Non - Responsive	The quoted model does not meet the following tender specification: i. It should have a dual temperature zone. ii. Should have Triple-pane windows with heat reflective film to reduce heat. iii. Should have 4 casters with 2 adjustable leveling feet.
3.	M/s. Gitansh Enterprises Pvt. Ltd. Make: Rolex Scientific Engineers	L3	Substantially Responsive	-----

Sd/-

General Manager (E)(i/c)