

16.10.2023

Online Tender for the Supply and Installation of Transport Ventilators To Three Tertiary Care Hospitals at Govt. Rajaji Hospital, Madurai, Govt. Kilpauk Medical College Hospital, Chennai and Govt. Coimbatore Medical College Hospital, Coimbatore

Ref: Tender No. PKG29(R)/C1/LCB/TNUHP/JICA/TNMSC/ENGG/2023,

Dt. 10.05.2023

Result of Evaluation

Sl. No.	Name of the bidder with make and model	Responsiveness		Remarks/ Reason for Rejection
		Technical Responsiveness	Commercial Responsiveness	
1.	M/s. Transhealthcare Make: Modul group, Turkey Model: Oxivent Oxi-4 Plus	Responsive	Responsive	-----
2.	Ms/. Medifocus India Pvt. Ltd. Make: Philips Model:Respironics Triology Evo	Non - Responsive	Responsive	As per the specialist report, the quoted model does not meet the following tender specification: 3. (b) Pressure vs Time - <i>Curve not available.</i> (c) Flow Vs Time - <i>Curve not available</i> (d) Expired values such as MV, total rate, MV spontaneous, FiO2, Vte, PEEP, Pmean, Peak Pr, with integrated EtCO2 - <i>EtCO2 external module</i>
3.	M/s. MGM Associates Make: Weinmann GmbH Model: Meduvent Standard with VC & PC	Non - Responsive	Responsive	As per the specialist report, the quoted model does not meet the following tender specification: 3. (b) Pressure vs Time - <i>Curve not available.</i> (c) Flow Vs Time - <i>Curve not available</i>

Sl. No.	Name of the bidder with make and model	Responsiveness		Remarks/ Reason for Rejection
		Technical Responsiveness	Commercial Responsiveness	
				<p>(d) Expired values such as MV, total rate, MV spontaneous, FiO₂, Vte, PEEP, Pmean, Peak Pr, with integrated EtCO₂ – <i>EtCO₂ not integrated.</i></p> <p>5 (b) BIPAP with / without pressure support – <i>not available</i></p> <p>7 (g) Oxygen high pressure hose – Not available</p>
4.	M/s. Solokrafts Industries Make: Air Liquide Model: Monnal T60	Responsive	Non - Responsive	Bidder does not have past experience for supply of similar model ventilator.

Sd/-

General Manager (E)(i/c)