

**Dt.16.10.2023**

**EVALUATION REPORT RESULT**

**Online Tender for Fixing Rate contract for supply of VDRL Rotator**

**Tender no.E948/VDRL/RC/TNMSC/ENGG/2023, Dt.30.03.2023**

Sl. No	Name of the Bidder	Responsiveness of bidder		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	<b>M/s. Alan Medical &amp; Laboratory Products,</b> Make: REMI, Model: RS-12 Plus	<b>Responsive</b>	<b>Responsive</b>	-----
2	<b>M/s. RL Healthcare,</b> Make: Rotek Plus, Model: PVD-R01 D	Non-Responsive	Non-Responsive	<p>1) The quoted model did not meet the following technical points:-</p> <ul style="list-style-type: none"> <li>• Should be suitable for wide variety of rotating and mixing applications.</li> <li>• Should have a maximum speed of 180 RPM and suitable for serological, VDRL tests and other tuned clinical diagnostic procedures.</li> <li>• Should have spring holders for flasks and blood bottles.</li> </ul> <p>2) Despite given chance did not furnish the documents such as Manufacturer authorization &amp; capacity, annual turnover, cash credit limit, profit loss statement and balance sheet. Also not furnished the end user certificates with supporting documents like purchase order and installation report for the supplies made by them as a non-manufacturer and also for their manufacturer supply in any one of the last five years along with the affidavit.</p>
3	<b>M/s. S.M. Scientific Instruments Pvt. Ltd,</b> Make: Jindal, Model: SMI-209	Non-Responsive	Responsive	<p>The quoted model did not meet the following technical points:-</p> <ul style="list-style-type: none"> <li>• Should be suitable for wide variety of rotating and mixing applications.</li> <li>• Should have a maximum speed of 180 RPM and suitable for serological, VDRL tests and other tuned clinical diagnostic procedures.</li> <li>• Should have spring holders for flasks and blood bottles.</li> </ul>