

**ONLINE TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND
INSTALLATION OF PORTABLE VENTILATOR TO VARIOUS GOVT.
INSTITUTIONS**

Tender No. E918/VEN/RC/TNMSC/ENGG/2023, DT: 22.02.2023

Result of Evaluation

Sl. No	Name of the firm	Model & Make	Technical evaluation report	Commercial evaluation report	Reason for rejection
1.	M/s. Schiller Healthcare India Pvt. Ltd	Model: Flight 60/ Flight Medical Innovations Ltd., Israel	Responsive	Responsive	-----
2.	M/s. Medifocus India Pvt. Ltd.	Hamilton T1 / Hamilton Medical, Switzerland	Responsive	Non-Responsive	<p>A. Despite given chance the bidder did not furnish documentary evidence meeting the following terms of the tender as a non-manufacturer bidder:-</p> <p>1. The bidder, as authorized by the manufacturer, must have supplied and provided after sales service to the extent of at least 1 No. for Sl. No.1 of the quantities of similar model in the Schedule of Requirements in any one of the last five (5) years to Indian Institution, which must be in satisfactory operation at least for one year on the date of bid opening. Hence non-responsive.</p>
3.	M/s. Trans Healthcare India pvt. Ltd.	Oxivent Oxi4plus / Modul Group Muhendislik Ltd., STI, Turkey	Non-Responsive	Responsive	<p>As per the demo report following technical specification of the tender are not met by the quoted model:-</p> <p>1. Secondary internal backup minimum of 30 minutes.</p> <p>2. Airborne Certified DO-160FIEC 60601-1, IEC 60601-1-2, IEC 60601-2-12, ASTM 1246 F etc. for transport condition. Hence not recommended by the specialists.</p>

Sl. No	Name of the firm	Model & Make	Technical evaluation report	Commercial evaluation report	Reason for rejection
4.	M/s. ResMed India Pvt. Ltd	ResMed Astral 150	Non-Responsive	Responsive	<p>As per the demo report following technical specification of the tender are not met by the quoted model:-</p> <ol style="list-style-type: none"> 1. PV and FV loops. 2. VCV starts only from 50 ml. against the requirement of 30 ml. 3. Advance modes namely (PRVC/ volume guarantee/ equivalent mode, MMV/ Minute volume guarantee/ ASV mode). 4. Secondary internal backup minimum of 30 minutes. 5. Advance features like adjustable slope rise, Adjustable expiratory threshold, automatic altitude compensations. 6. Respiratory mechanics like resistance & Compliance (static and Dynamic), RSBI Auto Beep. 7. Nebulizer port volume compensated. 8. Adjustable Expiratory Threshold. 9. Measurement of resistance and compliance. 10. Airborne Certified DO-160FIEC 60601-1 <p>Hence not recommended by the specialists.</p>
5.	M/s. Cnergy Orthocare (P) Ltd.	Shangrila 510S / Aeonmed, China	Non-Responsive	Non-Responsive	<p>A. As per the demo report following technical specification of the tender are not met by the quoted model:-</p> <ol style="list-style-type: none"> 1. 6" screen (instead only 5" screen). 2. Bilevel or BiPAP or equivalent. 3. Advance modes VIZ., (PRVC/ volume guarantee/ equivalent mode, MMV/ Minute volume guarantee/ ASV mode). 4. VCV requirement of 30 ml.

Sl. No	Name of the firm	Model & Make	Technical evaluation report	Commercial evaluation report	Reason for rejection
					<p>(Instead starts from 50ml).</p> <p>5. Secondary internal backup minimum of 30.</p> <p>6. Advance features like adjustable slope rise, Adjustable expiratory threshold, automatic altitude compensations.</p> <p>7. Facility to records up to 1000 events with parameter setting.</p> <p>8. Nebulizer port volume compensated.</p> <p>9. Adjustable Expiratory Threshold.</p> <p>10.Measurement of resistance and compliance.</p> <p>11.Airborne Certified DO-160FIEC 60601-1</p> <p>B. Despite given chance the bidder did not furnish documentary evidence meeting the following terms of the tender as a non-manufacturer bidder:-</p> <p>1. The bidder, as authorized by the manufacturer, must have supplied and provided after sales service to the extent of at least 1 No. for Sl. No.1 of the quantities of similar model in the Schedule of Requirements in any one of the last five (5) years to Indian Institution, which must be in satisfactory operation at least for one year on the date of bid opening.</p> <p>2. Documentary evidence (end users certificate) in support of the satisfactory operation of the goods are less than one year.</p> <p>Hence non-responsive.</p>

Sl. No	Name of the firm	Model & Make	Technical evaluation report	Commercial evaluation report	Reason for rejection
6.	M/s. Solokrafts Industries	Monnal T60 / Air Liquide Medical Systems SA, France	Non-Responsive	Non-Responsive	<p>A. As per the demo report following technical specification of the tender are not available:-</p> <ol style="list-style-type: none"> 1. Battery backup 2.30 hours only against 3 hours requirement. 2. Pressure support ranges from 5 to 60 cm H₂O, against 0 to 50 cm H₂O. 3. Nebulizer port volume compensated. <p>Hence not recommended by the specialists.</p> <p>B. Despite given chance the bidder did not furnish documentary evidence meeting the following terms of the tender as a non-manufacturer bidder:-</p> <ol style="list-style-type: none"> 1. The bidder, as authorized by the manufacturer, must have supplied and provided after sales service to the extent of at least 1 No. for Sl. No.1 of the quantities of similar model in the Schedule of Requirements in any one of the last five (5) years to Indian Institution, which must be in satisfactory operation at least for one year on the date of bid opening. <p>Hence non-responsive.</p>