

Dt.22.11.2023

**EVALUATION REPORT RESULT****Tender no.E1027/HAO/NHM//TNMSC/ENGG/2023, Dt.26.07.2023****Online Tender for Fixing Rate contract for supply and installation of Hot Air Oven**

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	<b>M/s. Asean Scientific Express, Chennai.</b> Make: Original Equipment Manufacturer, Model: ASE-HO3	<b>Responsive</b>	<b>Responsive</b>	-----
2	<b>M/s. Digiquel Systems, Chennai</b> Make: Original Equipment Manufacturer	<b>Responsive</b>	<b>Responsive</b>	-----
3	<b>M/s. York Scientific Industries Pvt. Ltd, Ghaziabad.</b> Make: Original Equipment Manufacturer, Model: YSI-431D	<b>Responsive</b>	<b>Responsive</b>	-----
4	<b>M/s. Madras Healthcare, Chennai.</b> Make: Saini Scientific Industry	Responsive	Non-responsive	<ul style="list-style-type: none"> <li>➤ Despite given chance did not furnish the enduser certificate along with supporting documents like purchase order, installation report to meet the qualification criteria.</li> <li>➤ Also not furnished the Annual turnover, cash credit limit, Profit &amp; Loss Statement, Balance Sheet and manufacturing capacity as per the bid document. Hence non-responsive in commercial evaluation.</li> </ul>
5	<b>M/s. Krishna Scientific Suppliers, Chennai</b> Make: Original Equipment Manufacturer	Non-responsive	Non-responsive	<ul style="list-style-type: none"> <li>➤ Against the technical point "Facility for adjustable shelves, 3 removable shelves to be provided" – Offered only 2 shelves.</li> </ul>

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
				<p>Hence non-responsive in technical evaluation.</p> <p>➤ Despite given chance did not furnish the enduser certificate along with supporting documents like purchase order, installation report to meet the qualification criteria. Hence non-responsive in commercial evaluation.</p>
6	<b>M/s. Gitansh Enterprises, Delhi.</b> Make: Rolex, Model: SAIH-HAO-30	Non-responsive	Non-responsive	<p>➤ As per the manufacturer authorization, <u>M/s. Rolex Scientific Engineers, Haryana</u> is the manufacturer of Hot Air Oven has authorized M/s. Gitansh Enterprises Pvt. Ltd to bid for this tender. However in the technical bid model is mentioned as "SAIH-HAO-30" and in the brochure make is mentioned as "SPIRE". Hence non-responsive in technical evaluation.</p> <p>➤ M/s. Gitansh Enterprises Pvt. Ltd is incorporated on 13.01.2022.</p> <p>➤ Further they did not furnish the enduser certificate along with supporting documents like purchase order, installation report to meet the qualification criteria. The documents submitted are only for Hydrotherapy items supplied to IIYNMS, Chengalpattu. Also not</p>

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
				furnished the Annual turnover, cash credit limit, Profit & Loss Statement, Balance Sheet and manufacturing capacity as per the bid document. The documents submitted are in the name of M/s. Rolex Scientific Engineers, Haryana. Hence non-responsive in commercial evaluation.
7	<b>M/s. Servell Bio Engineers Pvt. Ltd, Chennai.</b> Make: Hasthas, Model: SE-HAS-125L	<u>Bid security not furnished.</u> Hence their bid is treated as non-responsive.		
8	<b>M/s. Zenotrex Medico Pvt. Ltd, Haryana.</b> Make: Original Equipment Manufacturer, Model: ZPL-H01O	<u>Bid security not furnished.</u> "Domestic SS Enterprises who are the manufacturers of quoted product within the State and registered with MSME are alone eligible for EMD exemption, otherwise not eligible". Hence their bid is treated as non-responsive.		