

Dt.28.11.2023

EVALUATION REPORT RESULT

Online Tender for Fixing Rate contract for supply and installation of Direct Ophthalmoscope, BIOM Re-Inverter and OCT (Low End) to various Govt. Institutions

Tender no.E1028/EYE/RC/TNMSC/ENGG/2023, Dt.26.07.2023

Sch.no.1 Direct Ophthalmoscope

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. South India Surgical Co. Ltd, Make: Keeler	Responsive	Responsive	-----
2	M/s. Gitansh Enterprises Pvt. Ltd, Make: Rolex	Non-Responsive	Non-Responsive	<ul style="list-style-type: none">➤ As per the manufacturer authorization, <u>M/s. Rolex Scientific Engineers, Haryana</u> is the manufacturer of Direct Ophthalmoscope, BIOM Re-Inverter and OCT (Low End) has authorized M/s. Gitansh Enterprises Pvt. Ltd to bid for this tender. However during demonstration, the bidder has demonstrated different make of Direct Ophthalmoscope of <u>Heine, Germany.</u> Hence not recommended.➤ M/s. Gitansh Enterprises Pvt. Ltd is incorporated on 13.01.2022.➤ Further they did not furnish the enduser certificate along with supporting documents like purchase order, installation report to meet the qualification criteria. The documents submitted are only for Hydrotherapy items supplied to IIYNMS, Chengalpattu.

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
				➤ Also not furnished the Annual turnover, cash credit limit, Profit & Loss Statement and Balance Sheet as per the bid document. The documents submitted are in the name of M/s. Rolex Scientific Engineers, Haryana. Hence not recommended.

Sch.no.2 BIOM Re-Inverter

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Asian Vision, Make: Icon Medicare, Model: Reti view	Responsive	Responsive	-----
2	M/s. Appasamy Associates Pvt. Ltd, Make: Appasamy	Non-responsive	Responsive	➤ Against the technical point "Sterilization cases" offered "ETO case only". Hence not recommended.
3	M/s. Gitansh Enterprises Pvt. Ltd, Make: Rolex	Non-responsive	Non-responsive	➤ Not demonstrated. ➤ M/s. Gitansh Enterprises Pvt. Ltd is incorporated on 13.01.2022. ➤ Further they did not furnish the enduser certificate along with supporting documents like purchase order, installation report to meet the qualification criteria. The documents submitted are only for Hydrotherapy items supplied to IIYNMS, Chengalpattu. ➤ Also not furnished the Annual turnover, cash credit limit, Profit & Loss Statement and Balance Sheet as per the bid document. The documents submitted are in the name of M/s. Rolex Scientific Engineers, Haryana. Hence not recommended.

Sch.no.3 OCT (Low End)

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Appasamy Associates Pvt. Ltd, Make: Appasamy, Model: Appa OCT Ciana	Responsive	Responsive	-----
2	M/s. Gitansh Enterprises Pvt. Ltd, Make: Rolex	Non-responsive	Non-responsive	<ul style="list-style-type: none"> ➤ Not demonstrated. ➤ M/s. Gitansh Enterprises Pvt. Ltd is incorporated on 13.01.2022. ➤ Further they did not furnish the enduser certificate along with supporting documents like purchase order, installation report to meet the qualification criteria. The documents submitted are only for Hydrotherapy items supplied to IIYNMS, Chengalpattu. ➤ Also not furnished the Annual turnover, cash credit limit, Profit & Loss Statement and Balance Sheet as per the bid document. The documents submitted are in the name of M/s. Rolex Scientific Engineers, Haryana. Hence not recommended.