

Dt.28.11.2023

EVALUATION REPORT RESULT

**Online Tender for Fixing Rate Contract for supply and installation of
Tonopen, Motorized Ophthalmic Operation Table, B Scan, A Scan and Non-
Contact Tonometer to various Govt. Institutions**

Tender no.E1018/EYE/RC/TNMSC/ENGG/2023, Dt.30.06.2023

Sch.no.2 Motorized Ophthalmic Operation Table

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Appasamy Associates Pvt. Ltd, Make: Appasamy, Model: AAOT	Responsive	Responsive	-----
2	M/s. South India Surgical Co. Ltd, Make: Surgical Equipment Co, Model: SECO-1300E	Responsive	Responsive	-----

Sch.no.3 B Scan

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Appasamy Associates Pvt. Ltd, Make: Appasamy, Model: Marvel	Responsive	Responsive	-----
2	M/s. Punith Enterprises, Make: Akas Technologies, Model: Explorer	Responsive	Responsive	-----
3	M/s. Solokrafts Industries, Make: Rumax, Model: Matronix	Non-responsive	Non-responsive	<ul style="list-style-type: none">➤ The quoted make and model not meet the following technical point:-<ul style="list-style-type: none">• Touch Screen user interface – not available.➤ Despite given chance, did not furnish the enduser certificate to meet the qualification criteria with documentary evidence like purchase order, installation report etc for manufacturer supply.

Sch.no.4 A Scan

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Appasamy Associates Pvt. Ltd, Make: Appasamy, Model: Appascan-Max	Responsive	Responsive	-----
2	M/s. Punith Enterprises, Make: Akas Technologies, Model: Bioscale	Responsive	Responsive	-----
3	M/s. Solokrafts Industries, Make: Rumax, Model: i-Tronix	Non-responsive	Non-responsive	<p>➤ The quoted make and model not meet the following technical point:-</p> <ul style="list-style-type: none"> • <u>Formulas</u> - Post Refractive: Latkany Myopic, Latkany Hyperop, Aramberri Double-K, Comparative IOL calculations for upto 8 Lenses – not available. <p>➤ Despite given chance, did not furnish the enduser certificate to meet the qualification criteria with documentary evidence like purchase order, installation report etc for manufacturer supply.</p>

Sch.no.5 Non Contact Tonometer

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Appasamy Associates Pvt. Ltd, Make: Canon Singapore Pte. Ltd, Model: TX-20P	Responsive	Responsive	-----