

06.12.2023

**ONLINE TENDER FOR SUPPLY AND INSTALLATION OF FURNITURE ITEMS-I TO THREE
TERTIARY CARE HOSPITAL AT GOVT. RAJAJI HOSPITAL, MADURAI; GOVT KILPAUK
MEDICAL COLLEGE HOSPITAL, CHENNAI AND GOVT. COIMBATORE MEDICAL COLLEGE
HOSPITAL, COIMBATORE PKG57/C1.2/P2/LCB/TNUHP/JICA/TNMSC/ENGG/2023, Dt.**

09.03.2023

RESULT OF EVALUATION

Sch. No.	Name of the Bidder	Responsiveness on Technical Points	Responsiveness on Commercial Terms	Reason for rejection/ Remarks
I	Mattress with Pillow – 15 nos.			
1.	M/s. Medifocus India Pvt. Ltd., Chennai Make:- Drive DeVilbiss Healthcare Model :- MED AIRE Flexwave System	Responsive	Responsive	-----
2.	M/s. Lakshmi Agency Make:- Alerta Ruby Auto Model:- ALT-9009	Non Responsive	Responsive	<u>Technical point:</u> 5) Should automatic button lock after 30 seconds – Available with 5 minutes 12) Weight : 3.6 Kg – Available with 10 kgs Hence not recommended
3.	M/s. Carewell Medical Systems Make:- POCT (Point of Care) Model:- PAB-10	Non Responsive	Responsive	<u>Technical point:</u> Product not available with the bidder as per the Technical Corrigendum dated 26.06.2023. Hence not recommended
II	Stretcher Bed 3 Position IV Pole with Mattress – 5 nos.			
1.	M/s. Solokrafts Industries Make:- Own Model:- Kraft 401A	Responsive	Responsive	-----
2.	M/s. Imedfurns Pvt. Ltd., Chennai Make :- IMEDFURNS Model :- IMED5601T	Responsive	Responsive	-----
III	Transport Stretcher with Oxygen Cylinder IV Stand – 113 nos.			
1.	M/s. Solokrafts Industries Make:- Own Model:- Kraft 402A	Responsive	Responsive	-----
2.	M/s. BHN Surgicals Pvt. Ltd. Make :- BHN Model :- MW-21(B)	Responsive	Responsive	-----
3	M/s. Imedfurns Pvt. Ltd., Chennai Make :- IMEDFURNS Model :- IMED5621SST	Responsive	Responsive	-----