

Result of Evaluation and Final Short Listed Bidder

Online tender for fixing rate contract for supply and installation of Treadmill testing machine to various Govt. Institution

Tender Ref. No. E1033/TMT/DMS/TNMSC/ENGG/2023, dt. 26.07.2023

Sch: - 1 – Treadmill Testing Machine – Low End

Sl. No	Name of the Bidders	Responsiveness		Reason for Rejection
		Technical evaluation	Commercial evaluation	
1.	M/s. Surya Surgical Industries Make:- Medicaid Systems Model:- Cardivision	Non – responsive	Non – responsive	<p>As per specialist report the bidder is not meeting the following technical points:-</p> <ol style="list-style-type: none">1. Amplifier - Sampling rate – 2700 samples / sec CMRR>100db.2. Pre Programmed protocols including Bruce, modified Bruce, Naughton, Balke, Naughton facility3. The final report includes information on Blood Pressure, Heart rate, Treadmill Speed / Grade, ST Trends relating to stage wise & recovery phase. Simultaneous acquisition & display of 12 lead Raw ECG during the resting & exercise ECG mode. Time based METS calculation. Should have facility of on line storage of patient ECG data. Storage of atleast 500 patients on HDD. The monitor should display auto comparison of resting versus current lead of maximum ST depression separately with color coded protocol, stage, clocks for elapsed time, total time, Target HR, Treadmill speed & grade, PVC counts / minute, warning messages & prompts, lead check torso.4. Sample speed of TMT 1000s/sec. <p>As per commercial evaluation report the bidder is not meeting the following commercial points:-</p> <ol style="list-style-type: none">1. Non- Manufacturer Bidders The bidder, as authorized by the manufacturer, must have supplied and provided after sales service to the extent of at least 1 No. for Sl. No.1 & 2 of the quantities of similar model in the Schedule of Requirements in any one of the last five (5) years to Indian Institution, which must be in satisfactory operation at least for one year on the date of bid opening.2. Affidavit confirming that the performance statement given is correct along with copy of purchase order, copy of invoices, proof of payment received from Purchasers, documentary evidence (end user certificate) in support of satisfactory completion of orders and function as stated above.

2.	M/s. Allengers global Healthcare Pvt. Ltd., Make: Allengers Global Healthcare Pvt. Ltd. Model:- Gemini-A-DX	Responsive	Responsive	-----
3.	M/s. Schiller Healthcare India Pvt. Ltd., Make:- Schiller Healthcare Model:- CARDIOVIT CS 10	Responsive	Responsive	-----

Sch: - 2 – Treadmill Testing Machine – High End

Sl. No	Name of the Bidders	Responsiveness		Reason for Rejection
		Technical evaluation	Commercial evaluation	
1.	M/s. Surya Surgical Industries Make:- Labtech Model:- EC - 12LT	Non – responsive	Non – responsive	<p>As per the specialist report</p> <p>1. The model demonstrated does not meet the specification.</p> <p>As per commercial evaluation report the bidder is not meeting the following commercial points:-</p> <p>1. Non- Manufacturer Bidders The bidder, as authorized by the manufacturer, must have supplied and provided after sales service to the extent of at least 1 No. for Sl. No.1 & 2 of the quantities of similar model in the Schedule of Requirements in any one of the last five (5) years to Indian Institution, which must be in satisfactory operation at least for one year on the date of bid opening.</p> <p>2. Affidavit confirming that the performance statement given is correct along with copy of purchase order, copy of invoices, proof of payment received from Purchasers, documentary evidence (end user certificate) in support of satisfactory completion of orders and function as stated above.</p>

2.	M/s. Schiller Healthcare India Pvt. Ltd., Make:- Schiller Healthcare Model:- Cardiovit CS 200 Excellence	Responsive	Responsive	-----
----	---	------------	------------	-------