

Result of Evaluation and Final Short Listed Bidder

**Online tender for fixing rate contract for supply and installation of
Anesthesia Workstation**

Tender Ref. No. E966/ANS/RC/TNMSC/ENGG/2023,dt.18.04.2023

Sch: - 1 – Anaesthesia Workstation Type – I

Sl. No	Name of the Bidders	Responsiveness		Reason for Rejection
		Technical evaluation	Commercial evaluation	
1.	M/s. Solokrafts Industries Make: Mindray Model: WATO EX -35	Responsive	Responsive	-----
2.	M/s. Skanray Technologies Pvt. Ltd., Make: Skanray Model: Skan Siesta	Non - Responsive	Responsive	As per specialist report the bidder is not meeting the following technical points:- 1. Vaporiser should be European CE or USFDA certified – Drager certificate is shown 2. Should have fresh gas coupled/decoupled semi closed absorber system – Not demonstrated properly. 3. Should have an integrated/external fresh gas outlet for connecting Magill or Bain's circuit with electronic detection on screen for added patient safety – ACGO is not displayed in screen. In ACGO mode ventilator continues to work.
3.	M/s. Madras Surgicals & Chemicals Make: Drager Model: Fabius plus XL	Responsive	Responsive	-----

Note: - For CSDCO import license M/s. Madras Surgicals & Chemicals, M/s. Skanray Technologies Pvt. Ltd. and M/s. Solokrafts Industries applied and furnished file. no.

Sch: - 2 – Anaesthesia Workstation Type – II

Sl. No	Name of the Bidders	Responsiveness		Reason for Rejection
		Technical evaluation	Commercial evaluation	
1.	M/s. Solokrafts Industries Make: Mindray Model: WATO EX -30	Responsive	Responsive	-----

2.	M/s. Skanray Technologies Pvt. Ltd., Make: Skanray Model: SkanSiesta	Non – responsive	Responsive	<p>As per specialist report the bidder is not meeting the following technical points:-</p> <ol style="list-style-type: none"> 1. Should have fresh gas coupled/decoupled semi closed absorber system – Not demonstrated properly. 2. Should have fresh gas de-coupled/ Fresh gas flow compensation semi closed circle absorber system – Not demonstrated properly. 3. Should have an integrated/external fresh gas outlet for connecting Magill or Bain's circuit or pediatric circuit – When switched to ACGO mode, ACGO is not displayed in screen. In ACGO mode for bain's circuit the ventilator continues to work which is not safe.
3.	M/s. Trans Health Care Pvt. Ltd., Make: Shenzen Comen Medical Instruments Co. Model: AX - 400	Responsive	Responsive	-----
4.	M/s. Cnergy Orthocare Pvt. Ltd., Make: Beijing Aeonmed Co., Ltd. Model: Aeon 8600A	Non – responsive	Non – responsive	<p>As per specialist report the bidder is not meeting the following technical points:-</p> <ol style="list-style-type: none"> 1. Should furnish CDSCO import licence certificate for the quoted model – Not furnished. 2. Dual Cascade type flow meter tube for Oxygen & N2O. Range 20ml/min or less to 10 Lit/min or more. Calibrated in multiple scales. Single or dual tube for air 100ml or less to 10 L/min or more – Digital flow meter cannot work in case of power failure & display failure. 3. Should have optimized absorber canister approx. 1.5 Ltr. The CO2 absorber should be reusable and Autoclavable and should be easily detachable from the system without interrupting during active ventilation CO2 bypass function – Not demonstrated only video shown. 4. It should have bypass function for changing soda lime intra-operatively without losing circuit pressure.

				<p>As per commercial evaluation report the bidder is not meeting the following commercial points:-</p> <p>1. Manufacturer Bidder The bidder must have manufactured and supplied satisfactorily the similar model quoted in each schedule of the Schedule of Requirements either directly or through of any other authorized dealer to the extent of at least 1 No. for all Sl. No. under "Section – V, for Schedule of Requirements" in any one of the last five years of which minimum 1 No. for all Sl. No. should have been supplied to Indian Institution and should be in use satisfactorily with no adverse report for at least one year preceding the date of bid opening.</p> <p>2. Non-Manufacturer Bidder The bidder, as authorized by the manufacturer, must have supplied and provided after sales service to the extent of at least 1 No. for all of the quantities of similar model in the Schedule of Requirements in any one of the last five (5) years to Indian Institution, which must be in satisfactory operation at least for one year on the date of bid opening.</p> <p>3. Affidavit confirming that the performance statement given is correct along with copy of purchase order, copy of invoices, proof of payment received from Purchasers, documentary evidence (end user certificate) in support of satisfactory completion of orders and function as stated above.</p>
5.	M/s. BPL Medical Technologies Pvt. Ltd., Make:Penlon Model:Prima 320	Responsive	Responsive	-----

Note: - For CSDCO import license M/s. Skanray Technologies Pvt. Ltd., M/s. Trans Health Care Pvt. Ltd., M/s. BPL Medical Technologies Pvt. Ltd., and M/s. Solokrafts Industries applied and furnished file. no.