

Result of Evaluation and Finally Short Listed Bidders							
TENDER REF.NO. E001 /R-Consumables/RC/TNMSC/MAINT/2023, Dt.30.05.2023							
ITEM No.	Brief Description	Name of the Bidders					
		M/s.Madras Health Care		M/s.Kalyani radiotherapy Specialty(I) Pvt.Ltd.,		M/s.Meditronix	
		Technical Evaluation report	Commercial Evaluation Report	Technical Evaluation report	Commercial Evaluation Report	Technical Evaluation report	Commercial Evaluation Report
1	Thermoplastic mould compatible for existing All-in-One baseplate Head and Neck region 5 clamp-2.4mm thickness, 22% perforation	Not responsive. As per technical committee This product is not rigid and cannot be used for average/well-built patients as mold gets thinner.	Responsive	Not responsive.As per technical committee This product is not rigid and cannot be used for average/well-built patients.	Responsive	Responsive	Responsive
2	Thorax/Breast thermoplastic cast compatible with All-in-One base plate with 2.4mm thickness, 22%perforation	Not responsive. As per technical committee Mold length not adequate.	Responsive	Not responsive. As per technical committee 1) Sample melts after 1st use itself 2) Mold splitting during transport and hence no durable.	Responsive	Not responsive. As per technical committee Sample given is unfinished.	Responsive
3	Pelvic thermoplastic mask compatible with All-in-One base plate with 2.4mm thickness, 22%perforation	Not responsive. As per technical committee Sample is sticky and moldability inadequate for patient use.	Responsive	Not responsive.As per technical committee Sample moldability is inadequate	Responsive	Not responsive. As per technical committee Sample given is unfinished.	Responsive
4	Thermoplastic mould S type compatible for existing Q-fix make baseplate Head & Neck region 9 PIN-2.4mm thickness,22%perforation	Not responsive.As per technical committee This product is not rigid and cannot be used for average/well-built patients as mold gets thinner.	Responsive	Not responsive.As per technical committee This product is not rigid and cannot be used for average/well-built patients.	Responsive	Responsive	Responsive
5	Breast/ Thoracic thermoplastic cast compatible with Qfix breast board with 2.4mm thickness, 22%perforation	Not responsive. As per technical committee Mold length not adequate.	Responsive	Not responsive. As per technical committee 1) Sample melts after 1st use itself 2) Mold splitting during transport and hence no durable.	Responsive	Not responsive. As per technical committee Sample given is unfinished.	Responsive

<u>Result of Evaluation and Finally Short Listed Bidders</u>							
<u>TENDER REF.NO. E001 /R-Consumables/RC/TNMSC/MAINT/2023, Dt.30.05.2023</u>							
ITEM No.	Brief Description	Name of the Bidders					
		M/s.Madras Health Care		M/s.Kalyani radiotherapy Specialty(I) Pvt.Ltd.,		M/s.Meditronix	
		Technical Evaluation report	Commercial Evaluation Report	Technical Evaluation report	Commercial Evaluation Report	Technical Evaluation report	Commercial Evaluation Report
6	Pelvic thermoplastic mask compatible with Qfix portrait base plate with 2.4mm thickness, 22%perforation	Not responsive. As per technical committee Sample is sticky and moldability inadequate for patient use.	Responsive	Not responsive.As per technical committee Sample moldability is inadequate	Responsive	Not responsive. As per technical committee Sample given is unfinished.	Responsive
7	Carbon fiber Head rest (Type A) compatible with AIO / Qfix baseplate	Not responsive. As per technical committee Sample submitted is not carbon fiber and also not available in catalogue.	Responsive	Responsive	Responsive	Responsive	Responsive
8	Carbon fiber Head rest (Type B) compatible with AIO / Qfix baseplate	Not responsive. As per technical committee Sample submitted is not carbon fiber and also not available in catalogue.	Responsive	Responsive	Responsive	Responsive	Responsive
9	Carbon fiber Head rest (Type C) compatible with AIO / Qfix baseplate	Not responsive. As per technical committee Sample submitted is not carbon fiber and also not available in catalogue.	Responsive	Responsive	Responsive	Responsive	Responsive
10	Paediatric headrest compatible with Qfix baseplate	Not responsive. As per technical committee Sample not submitted	Responsive	Not responsive. As per technical committee Sample not submitted	Responsive	NA	NA
11	Head & Neck Thermoplastic mask for SRS/SRT treatment pairs of masks of reinforced material	Not responsive. As per technical committee Sample not submitted and also not available in catalogue.	Responsive	Not responsive. As per technical committee Sample is not reinforced	Responsive	Not responsive. As per technical committee Sample is not reinforced	Responsive
12	3.2mm thick Adapter to use with head and neck device made of carbon fibre and should have indexing markers on itself.	Responsive	Responsive	Responsive	Responsive	NA	NA
13	Bolus 30x30x1cm	Responsive	Responsive	Responsive	Responsive	Responsive	Responsive
14	Bolus 30x30x0.5cm	Responsive	Responsive	Responsive	Responsive	Responsive	Responsive

<u>Result of Evaluation and Finally Short Listed Bidders</u>							
<u>TENDER REF.NO. E001 /R-Consumables/RC/TNMSC/MAINT/2023, Dt.30.05.2023</u>							
ITEM No.	Brief Description	Name of the Bidders					
		M/s.Madras Health Care		M/s.Kalyani radiotherapy Specialty(I) Pvt.Ltd.,		M/s.Meditronix	
		Technical Evaluation report	Commercial Evaluation Report	Technical Evaluation report	Commercial Evaluation Report	Technical Evaluation report	Commercial Evaluation Report
15	Nylon fabric with waterproof lamination for treatment in radiotherapy size 100x70 (35 ltr)	Responsive	Responsive	Not responsive. As per technical committee Sample outer cover is thin and not durable.	Responsive	Not responsive. As per technical committee Sample outer cover is thin and not durable.	Responsive
16	Nylon fabric with waterproof lamination for treatment in radiotherapy size 100x220 (120 ltr)	Responsive	Responsive	Not responsive. As per technical committee Sample outer cover is thin and not durable.	Responsive	Not responsive. As per technical committee Sample outer cover is thin and not durable.	Responsive
17	Nylon fabric with waterproof lamination for treatment in radiotherapy size 120x75 (35 ltr)	Responsive	Responsive	Not responsive. As per technical committee Sample outer cover is thin and not durable.	Responsive	Not responsive. As per technical committee Sample outer cover is thin and not durable.	Responsive
18	Dental wax Box (1mm strips)	Not responsive. As per technical committee Sample not submitted	Responsive	Not responsive. As per technical committee Sample is not as per specification.	Responsive	Responsive	Responsive
19	Tegaderm 1624W Medium (Box of 100nos)	Responsive	Responsive	Responsive	Responsive	Not responsive. As per technical committee Sample not given.	Responsive
20	Markers fiducial - Lead silver coating (50 nos per box)	Responsive	Responsive	Responsive	Responsive	Not responsive. As per technical committee Sample not given.	Responsive
21	Breast Adapter (mold fix)	Not responsive. As per technical committee Sample not submitted	Responsive	Not responsive. As per technical committee Sample not submitted	Responsive	Not responsive. As per technical committee Sample not given.	Responsive
22	Pelvis Adapter (mold fix)	Not responsive. As per technical committee Sample not submitted	Responsive	Not responsive. As per technical committee Sample not submitted	Responsive	Not responsive. As per technical committee Sample not given.	Responsive
23	Indexer Bar for Vac loc-Standard size compatible with existing CT and teletherapy machine	Not responsive. As per technical committee Sample not submitted	Responsive	Not responsive. As per technical committee Sample not submitted	Responsive	Not responsive. As per technical committee Sample given is not as per technical specification.	Responsive
24	OPTI Handles for Thorax/Breast, Pelvis thermoplastic casts compatible with existing Q-fix base plate	Not responsive. As per technical committee Sample not submitted	Responsive	Not responsive. As per technical committee Sample not submitted	Responsive	NA	NA

<u>Result of Evaluation and Finally Short Listed Bidders</u>	
---	--

TENDER REF.NO. E001 /R-Consumables/RC/TNMSC/MAINT/2023, Dt.30.05.2023

ITEM No.	Brief Description	Name of the Bidders					
		M/s.Madras Health Care		M/s.Kalyani radiotherapy Specialty(I) Pvt.Ltd.,		M/s.Meditronix	
		Technical Evaluation report	Commercial Evaluation Report	Technical Evaluation report	Commercial Evaluation Report	Technical Evaluation report	Commercial Evaluation Report
25	Indexer Bar compatible with all in one base plate	Not responsive. As per technical committee Sample given not compatible with AIO board.	Responsive	Responsive	Responsive	NA	NA
Sd/- General Manager (Services)							