

Online tender for fixing rate contract for supply of Micropipette to  
various Govt. Institutions

Online tender ref. E911/PIP/RC/TNMSC/ENG/2023, Dt.05.01.2023

**EVALUATION REPORT**

Sch No.	Item Description	Name of the bidder	Responsive ness of the bidder		Reason for rejection
			Technical Evaluation report	Commercial evaluation report	
		Fixed Micropipette			
1.	Fixed Micropipette 1µl	M/s.Alan Medical & Laboratory Products., Make: RV Instruments Model: AC600680 Catalogue No. 600680	Non-responsive	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> (i) Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b>
		M/s.Shan Biotech & Diagnostics., Make: ThermoFisher Scientific Pvt.Ltd., Model: FINNPIPETTE F3 Catalogue No. 4650000-N	Responsive	Responsive	-----
		M/s.Eppendorf India Private Limited., Make: Eppendorf Model: Research Plus Catalogue No. 4925000014	Responsive	Responsive	-----
2.	Fixed Micropipette 5µl	M/s.Alan Medical & Laboratory Products., Make: RV Instruments Model: AC600720 Catalogue No. 600720	Non-responsive	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> • Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b>
		M/s.Shan Biotech & Diagnostics., Make: ThermoFisher Scientific Pvt.Ltd., Model: FINNPIPETTE F3	Responsive	Responsive	-----

Sch No.	Item Description	Name of the bidder	Responsive ness of the bidder		Reason for rejection
			Technical Evaluation report	Commercial evaluation report	
		Catalogue No. 4650010-N			
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 4925000030	Responsive	Responsive	-----
		<b>M/s.IKA India Pvt.Ltd.,</b> Make: IKA Model: PETTE fix 5µl Catalogue No. 20011219	Responsive	Responsive	-----
3.	Fixed Micropipette 10µl	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: AC600740 Catalogue No. 600740	<b>Non-responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> • Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b>
		<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: ThermoFisher Scientific Pvt.Ltd., Model: FINNPIPETTE F3 Catalogue No. 4650020-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3124000016	Responsive	Responsive	-----
		<b>M/s.IKA India Pvt.Ltd.,</b> Make: IKA Model: PETTE fix 10µl Catalogue No. 20011220	Responsive	Responsive	-----
4.	Fixed Micropipette 25µl	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: AC600780 Catalogue No. 600780	<b>Non responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b>

Sch No.	Item Description	Name of the bidder	Responsive ness of the bidder		Reason for rejection
			Technical Evaluation report	Commercial evaluation report	
					<ul style="list-style-type: none"> <li>Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b></li> </ul>
		<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: ThermoFisher Scientific Pvt.Ltd., Model: FINNPIPETTE F3 Catalogue No. 4650030-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3124000059	Responsive	Responsive	-----
		<b>M/s.IKA India Pvt.Ltd.,</b> Make: IKA Model: PETTE fix 25µl Catalogue No. 20011222	Responsive	Responsive	-----
5.	Fixed Micropipette 50µl	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: AC600800 Catalogue No. 600800	<b>Non responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> <ul style="list-style-type: none"> <li>Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b></li> </ul>
		<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: ThermoFisher Scientific Pvt.Ltd., Model: FINNPIPETTE F3 Catalogue No. 4650040-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3124000067	Responsive	Responsive	-----

Sch No.	Item Description	Name of the bidder	Responsive ness of the bidder		Reason for rejection
			Technical Evaluation report	Commercial evaluation report	
		<b>M/s.IKA India Pvt.Ltd.,</b> Make: IKA Model: PETTE fix 50µl Catalogue No. 20011223	Responsive	Responsive	-----
6.	Fixed Micropipette 100µl	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: AC600820 Catalogue No. 600820	<b>Non responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> • Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b>
		<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: ThermoFisher Scientific Pvt.Ltd., Model: FINNPIPETTE F3 Catalogue No. 4650050-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3124000075	Responsive	Responsive	-----
		<b>M/s.IKA India Pvt.Ltd.,</b> Make: IKA Model: PETTE fix 100µl Catalogue No. 20011224	Responsive	Responsive	-----
7.	Fixed Micropipette 200µl	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: AC600840 Catalogue No. 600840	<b>Non responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> • Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b>

Sch No.	Item Description	Name of the bidder	Responsive ness of the bidder		Reason for rejection
			Technical Evaluation report	Commercial evaluation report	
		<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: ThermoFisher Scientific Pvt.Ltd., Model: FINNPIPETTE F3 Catalogue No. 4650100-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3124000083	Responsive	Responsive	-----
		<b>M/s.IKA India Pvt.Ltd.,</b> Make: IKA Model: PETTE fix 200µl Catalogue No. 20011225	Responsive	Responsive	-----
8.	Fixed Micropipette 250µl	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: AC600860 Catalogue No. 600860	<b>Non responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> • Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b>
		<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: ThermoFisher Scientific Pvt.Ltd., Model: FINNPIPETTE F3 Catalogue No. 4650060-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3124000105	Responsive	Responsive	-----

Sch No.	Item Description	Name of the bidder	Responsive ness of the bidder		Reason for rejection
			Technical Evaluation report	Commercial evaluation report	
		<b>M/s.IKA India Pvt.Ltd.,</b> Make: IKA Model: PETTE fix 250µl Catalogue No. 20011226	Responsive	Responsive	-----
9.	Fixed Micropipette 500µl	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: AC600880 Catalogue No. 600880	<b>Non responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> • Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b>
		<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: ThermoFisher Scientific Pvt.Ltd., Model: FINNPIPETTE F3 Catalogue No. 4650070-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3124000113	Responsive	Responsive	-----
		<b>M/s.IKA India Pvt.Ltd.,</b> Make: IKA Model: PETTE fix 500µl Catalogue No. 20011227	Responsive	Responsive	-----
10.	Fixed Micropipette 1000µl	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: AC600900 Catalogue No. 600900	<b>Non responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> • Durable tip cone materials made of PVDF plastic to provide excellent chemical

Sch No.	Item Description	Name of the bidder	Responsive ness of the bidder		Reason for rejection
			Technical Evaluation report	Commercial evaluation report	
					resistance - <b>The tip cone is not PVDF material.</b>
		<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: ThermoFisher Scientific Pvt.Ltd., Model: FINNPIPETTE F3 Catalogue No. 4650080-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3124000121	Responsive	Responsive	-----
		<b>M/s.IKA India Pvt.Ltd.,</b> Make: IKA Model: PETTE fix 1000µl Catalogue No. 20011228	Responsive	Responsive	-----
		<b>Variable Micropipette</b>			
11.	Variable Micropipette 1-10µl	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: AC600440 Catalogue No. 600440	<b>Non responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> <ul style="list-style-type: none"><li>Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b></li></ul>
		<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: ThermoFisher Scientific Pvt.Ltd., Model: FINNPIPETTE F3 Catalogue No. 4640010-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3123000225	Responsive	Responsive	-----

Sch No.	Item Description	Name of the bidder	Responsive ness of the bidder		Reason for rejection
			Technical Evaluation report	Commercial evaluation report	
		<b>M/s.IKA India Pvt.Ltd.,</b> Make: IKA Model: PETTE Vario 0.5-10µl Catalogue No. 20011211	Responsive	Responsive	-----
12.	Variable Micropipette 2-20µl	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: AC600460 Catalogue No. 600460	<b>Non responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> • Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b>
		<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: ThermoFisher Scientific Pvt.Ltd., Model: FINNPIPETTE F3 Catalogue No. 4640030-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3123000233	Responsive	Responsive	-----
		<b>M/s.IKA India Pvt.Ltd.,</b> Make: IKA Model: PETTE Vario 2-20µl Catalogue No. 20011213	Responsive	Responsive	-----
13.	Variable Micropipette 5-50µl	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: AC600500 Catalogue No. 600500	<b>Non responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> • Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b>



Sch No.	Item Description	Name of the bidder	Responsive ness of the bidder		Reason for rejection
			Technical Evaluation report	Commercial evaluation report	
		<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: ThermoFisher Scientific Pvt.Ltd., Model: FINNPIPETTE F3 Catalogue No. 4640090-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3123000241	Responsive	Responsive	-----
14.	Variable Micropipette 20-200µl	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: AC600540 Catalogue No. 600540	<b>Non responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> • Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b>
		<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: ThermoFisher Scientific Pvt.Ltd., Model: FINNPIPETTE F3 Catalogue No. 4640050-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3123000250	Responsive	Responsive	-----
		<b>M/s.IKA India Pvt.Ltd.,</b> Make: IKA Model: PETTE Vario 10-100µl Catalogue No. 20011214	Responsive	Responsive	-----
15.	Variable Micropipette 10-100µl	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: AC600520 Catalogue No. 600520	<b>Non responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b>

Sch No.	Item Description	Name of the bidder	Responsive ness of the bidder		Reason for rejection
			Technical Evaluation report	Commercial evaluation report	
					<ul style="list-style-type: none"> <li>Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b></li> </ul>
		<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: ThermoFisher Scientific Pvt.Ltd., Model: FINNPIPETTE F3 Catalogue No. 4640040-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3123000241	Responsive	Responsive	-----
		<b>M/s.IKA India Pvt.Ltd.,</b> Make: IKA Model: PETTE Vario 20-200µl Catalogue No. 20011215	Responsive	Responsive	-----
16.	Variable Micropipette 100-1000µl	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: AC600620 Catalogue No. 600620	<b>Non-responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> <ul style="list-style-type: none"> <li>Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b></li> </ul>
		<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: ThermoFisher Scientific Pvt.Ltd., Model: FINNPIPETTE F3 Catalogue No. 4640060-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3123000268	Responsive	Responsive	-----

Sch No.	Item Description	Name of the bidder	Responsive ness of the bidder		Reason for rejection
			Technical Evaluation report	Commercial evaluation report	
		<b>M/s.IKA India Pvt.Ltd.,</b> Make: IKA Model: PETTE Vario 100-1000µl Catalogue No. 20011216	Responsive	Responsive	-----
<b>Multichannel (8-Ch) Micropipette without spring load technique</b>					
17.	Multichannel Micropipette 1-10µl	<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: Thermo Fisher Scientific Pvt.Ltd., Model: FINNPIPETTE F2 Catalogue No. 4662000-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3125000010	Responsive	Responsive	-----
		<b>M/s.IKA India Pvt.Ltd.,</b> Make: IKA Model: PETTE multi 8x0.5 - 10µl Catalogue No. 20017943	Responsive	Responsive	-----
		<b>M/s.Alan Medical and Laboratory Products.,</b>	<b>Non – Responsive</b>	<b>Non – Responsive</b>	Bidder withdrawn from the schedule, Hence not recommended.
18.	Multichannel Micropipette 5-50µl	<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: Thermo Fisher Scientific Pvt.Ltd., Model: FINNPIPETTE F2 Catalogue No. 4662010-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3125000036	Responsive	Responsive	-----
		<b>M/s.Alan Medical and Laboratory Products.,</b>	<b>Non – Responsive</b>	<b>Non – Responsive</b>	Bidder withdrawn from the schedule, Hence not recommended.

Sch No.	Item Description	Name of the bidder	Responsive ness of the bidder		Reason for rejection
			Technical Evaluation report	Commercial evaluation report	
19.	Multichannel Micropipette 10-100µl	<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: Thermo Fisher Scientific Pvt.Ltd., Model: FINNPIPETTE F2 Catalogue No. 4662020-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3125000036	Responsive	Responsive	-----
		<b>M/s.IKA India Pvt.Ltd.,</b> Make: IKA Model: PETTE multi 8x10- 100µl Catalogue No. 20017944	Responsive	Responsive	-----
		<b>M/s.Alan Medical and Laboratory Products.,</b>	<b>Non – Responsive</b>	<b>Non – Responsive</b>	Bidder withdrawn from the schedule, hence not recommended,
20.	Multichannel Micropipette 30-300µl	<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: Thermo Fisher Scientific Pvt.Ltd., Model: FINNPIPETTE F2 Catalogue No. 4662030-N	Responsive	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3125000052	Responsive	Responsive	-----
		<b>M/s.Alan Medical and Laboratory Products.,</b>	<b>Non – Responsive</b>	<b>Non – Responsive</b>	Bidder withdrawn from the schedule, Hence not recommended,
21.	Multichannel Micropipette 100-1000µl	<b>M/s.Shan Biotech &amp; Diagnostics.,</b> Make: Thermo Fisher Scientific Pvt.Ltd., Model: FINNPIPETTE NOVUS MULTICHANNEL PIPETTE Catalogue No. 46300800	Responsive	Responsive	-----

Sch No.	Item Description	Name of the bidder	Responsive ness of the bidder		Reason for rejection
			Technical Evaluation report	Commercial evaluation report	
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3125000214	Responsive	Responsive	-----
		<b>M/s.Alan Medical and Laboratory Products.,</b>	<b>Non – Responsive</b>	<b>Non – Responsive</b>	Bidder withdrawn from the schedule, hence not recommended,
<b>Multichannel (8-Ch) Micropipette with spring load technique</b>					
22.	Multichannel Micropipette 1-10µl with spring load technique	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: MC 060100 Catalogue No. 060100	<b>Non-responsive</b>	Responsive	-----
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3125000010	Responsive	Responsive	-----
23.	Multichannel Micropipette 5-50µl with spring load technique	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: MC 060120 Catalogue No. 060120	<b>Non-responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> • Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b>
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3125000036	Responsive	Responsive	-----
24.	Multichannel Micropipette 10-100µl with spring load technique	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: MC 060130 Catalogue No. 060130	<b>Non-responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> • Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip</b>

Sch No.	Item Description	Name of the bidder	Responsive ness of the bidder		Reason for rejection
			Technical Evaluation report	Commercial evaluation report	
					cone is not PVDF material.
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3125000036	Responsive	Responsive	-----
25.	Multichannel Micropipette 30-300µl with spring load technique	<b>M/s.Alan Medical &amp; Laboratory Products.,</b> Make: RV Instruments Model: MC 060150 Catalogue No. 060150	<b>Non-responsive</b>	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. <b>Hence not recommended.</b> • Durable tip cone materials made of PVDF plastic to provide excellent chemical resistance - <b>The tip cone is not PVDF material.</b>
		<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3125000052	Responsive	Responsive	-----
26.	Multichannel Micropipette 100-1000µl with spring load technique	<b>M/s.Eppendorf India Private Limited.,</b> Make: Eppendorf Model: Research Plus Catalogue No. 3125000214	Responsive	Responsive	-----
		<b>M/s .Alan Medical and Laboratory Products.,</b>	<b>Non – Responsive</b>	<b>Non – Responsive</b>	Bidder withdrawn from the schedule, Hence not recommended,

**Sd/-**  
**Managing Director**