

**RESULT OF EVALUATION AND FINALLY SHORTLISTED BIDDERS**

**ONLINE TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND  
INSTALLATION OF STAINLESS STEEL TOP COT AND ICU COT TO  
VARIOUS GOVT. INSTITUTIONS**

**TENDER NO. E1019/COT/RC/TNMSC/ENGG/2023, DT:30.06.2023**

**Schedule (1): Stainless Steel Cot**

Sl. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
1	M/s. South India Surgical Co. Ltd	Bio Breeze Medicare Engineering Solutions	Responsive	Responsive	-----
2	M/s. R.Madhusudhanan	Own make	Responsive	Responsive	-----
3	M/s. Solokrafts Industries	Kraft 151	Responsive	Responsive	-----
4	M/s. Godrej and Boyce Manufacturing Company Limited	Ayush Lite Sheet	Non Responsive	Responsive	<b>Technical</b> The quoted and demonstrated model is MS powder coated instead tender requirement of Stainless Steel cot.
5	M/s. Prince Hospital Furniture (Mfg) Co.,	PHF014	Responsive	Responsive	-----
6	M/s. Samar Steel Industries	SSI-HF104	Responsive	Responsive	-----
7	M/s. Premier Hospital Steel Furniture Mfg Co	PHSF010	Responsive	Responsive	-----
8	M/s. Imedfurns Private Limited	IMED6001B	Responsive	Responsive	-----
9	M/s. Jade Life Sciences Pvt Ltd	SK scientific Surgicals/ SKSS809E	Responsive	Non Responsive	<b>Commercial</b> The documents namely balance sheet, P.O copy of previous supply are in the name of M/s. Scientific & Surgicals are not in the bidder's name.
10	M/s. Carewell Medical Systems	CWMS/ CMS117	Non responsive	Responsive	<b>Technical</b>

Sl. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
					As per the specialist report the model was not demonstrated.

**Schedule (2): ICU Cot (Type-I)**

Sl. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
1	M/s. Blue Med Surgicals	Paramount	Responsive	Responsive	-----
2	M/s. Solokrafts Industries	Kraft 102 M/M	Responsive	Responsive	-----
3	M/s. Godrej and Boyce Manufacturing Company Limited	ACURA	Non Responsive	Responsive	<u>Technical</u> As per the specialist report the model is not having Manual operation mode. It has only electrical mode with battery backup.
4	M/s. MM Medical Systems	Morak Manufacturing /EV 500	Responsive	Responsive	-----
5	M/s. Prince Hospital Furniture (Mfg) Co.,	PHF 004	Responsive	Responsive	-----
6	M/s. Imedfurns Private Limited	IMED5103E	Responsive	Responsive	-----
7	M/s. Jade Life Sciences Pvt Ltd	SK scientific Surgical/ SKSS801E	Non-Responsive	Non Responsive	<u>Technical</u> As per the specialist report the model is not having Manual operation mode. It has only electrical mode with battery backup. <u>Commercial</u> The documents namely balance sheet, P.O copy of previous supply are in the name of

Sl. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
					M/s. Scientific & Surgicals are not in the bidder's name.
8	M/s. Premier Hospital Steel Furniture Mfg Co	PHSF002	Responsive	Non Responsive	<u>Commercial</u> Despite given chance they could not furnish the documents like PO copy for previous supply, to meet the qualification criteria.
9	M/s. Carewell Medical Systems	CWMS/ CMS777	Non responsive	Responsive	<u>Technical</u> As per the specialist report the model was not demonstrated.
10	M/s. Confident Dental Equipments Pvt. Ltd	CME 18 AM	Responsive	Non Responsive	<u>Commercial</u> Despite given chance they could not furnish the documents like PO copy for previous supply, to meet the qualification criteria.
11	M/s. South India Surgical Co. Ltd	Bio Breeze Medicare Engineering Solutions	Responsive	Non Responsive	<u>Commercial</u> Despite given chance they could not furnish the documents like PO copy for previous supply, to meet the qualification criteria.

**Schedule (3): ICU Cot – Type-II**

Sl. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
1	M/s. Blue Med Surgicals	Paramount	Responsive	Responsive	-----
2	M/s. Solokrafts Industries	Kraft 100 EXP	Responsive	Responsive	-----
3	M/s. Godrej and Boyce Manufacturing Company Limited	Acura Neo	Responsive	Responsive	-----
4	M/s. Prince Hospital Furniture (Mfg) Co.,	PHF004	Responsive	Responsive	-----
5	M/s. Imedfurns Private Limited	IMEDC8SB	Responsive	Responsive	-----
6	M/s. Jade Life Sciences Pvt Ltd	SK scientific Surgicals/ SKSS800	Responsive	Non Responsive	<u>Commercial</u> The documents namely balance sheet, P.O copy of previous supply are in the name of M/s. Scientific & Surgicals are not in the bidder's name.
7	M/s. Premier Hospital Steel Furniture Mfg Co	PHSF001	Responsive	Non Responsive	<u>Commercial</u> Despite given chance they could not furnish the documents like PO copy for previous supply, to meet the qualification criteria.
8	M/s. Carewell Medical Systems	CWMS/ CMS779	Non Responsive	Responsive	<u>Technical</u> As per the specialist report the model was not demonstrated.
9	M/s. Confident Dental Equipments Pvt. Ltd	ICU-4-042	Responsive	Non responsive	<u>Commercial</u> Despite given chance they could not furnish the documents like PO copy for previous supply, to meet the qualification criteria.
10	M/s. Samar Steel Industries	SSI-HF 101B	Non-Responsive	Non Responsive	<u>Technical</u>

SI. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
					As per the specialist report the model was not demonstrated.